



K090443  
P14

# LARSEN & TOUBRO LIMITED

ELECTRICAL & ELECTRONICS DIVISION - ELECTRONIC PRODUCTS

Mysore Campus, KIADB Industrial Area, Hebbal - Hootagalli, Mysore - 570 018 • Tel: +91 (821) 2405000 • Fax: +91 (821) 2402468

E-Mail

Ref.

JUN - 5 2009

Date: 05.01.2009

Page: 01 of 04

## 510(K) SUMMARY (Per section 807.92 (c))

<b>CONTACT DATA</b>			
<b>Submitter's Name</b>		Larsen & Toubro Limited	
<b>Address</b>		KIADB Industrial Area, Hebbal Hootagalli, Mysore - 570018, Karnataka, INDIA	
<b>Telephone</b>	91-821-2405439	<b>Fax</b>	91-821-2402468
<b>Contact Person</b>	A.B.Deshpande	<b>Title</b>	Head - Quality Assurance & Management Representative
<b>E-Mail address</b>		DeshpandeAB@myw.tlindia.com	
<b>Date the summary was prepared</b>		05.01.2009	



K090443

p.2/4

# LARSEN & TOUBRO LIMITED

ELECTRICAL & ELECTRONICS DIVISION - ELECTRONIC PRODUCTS

Mysore Campus, KIADB Industrial Area, Hebbal - Hoctagalli, Mysore - 570 018 • Tel: +91 (821) 2405000 • Fax: +91 (821) 2402468

E-Mail:

Ref:

Date: 05.01.2009

Page: 02 of 04

DEVICE	
Trade name	PLANET 55 Model 100
Common name	Patient Monitoring System
Classification name	Vital Signs Monitor

PREDICATE DEVICE IDENTIFICATION			
CFR21 Section	870.1025	Product code	MIX
Classification panel	Cardiovascular		
Device Class	Class II		
Legally marketed Comparison Device / K#	<ul style="list-style-type: none"> <li>PLANET 55 Patient Monitoring System (L&amp;T Medical Equipments &amp; Systems) / K071472</li> <li>Passport 2 Vital signs monitor with View 12 ECG Analysis Module (Datascopie Corp) / K020550</li> </ul>		

2d



K090443  
p314

# LARSEN & TOUBRO LIMITED

ELECTRICAL & ELECTRONICS DIVISION - ELECTRONIC PRODUCTS

Mysore Campus, KIADB Industrial Area, Hebbal - Hootagalli, Mysore - 570 018 • Tel: +91 (821) 2405000 • Fax: +91 (821) 2402468

E-Mail:

Ref:

Date: 05.01.2009

Page: 03 of 04

## DEVICE DESCRIPTION

Planet 55 Model 100 is a Multi-parameter patient monitoring system for continuous monitoring of the physiological parameter ECG (3/5 lead), Arrhythmia & ST analysis, Respiration, NIBP, Temperature, SpO2 and CO2.

PLANET 55 Model 100 is a 4-channel monitor with 8.4" TFT display capable of displaying ECG, Respiration, SpO2, CO2 digital values of HR/PR, SpO2, RR, Non-Invasive Blood Pressure (Systolic, Diastolic and Mean), Temperature, EtCO2, FiCO2 readings. It has selective 24/48/72 Hours tabular and graphical trends. It has special feature of NIBP having a trend of storing last 240 readings. It has Alarm Recall facility with last 24 patient alarms details. It has a two-channel thermal array recorder for printing of Tabular trends & waveforms. It has got optional communication features - USB, RS232, Infrared-remote and Ethernet. The device permits patient monitoring with adjustable alarm limits as well as visible and audible alarm signals.

## INTENDED USE OF THE DEVICE

The PLANET 55 Model 100 multi-parameter Patient Monitoring system is intended to monitor a single Adult, Pediatric or Neonatal patient's vital signs at the bedside or during intra-hospital transport along with the appropriate accessories mentioned / supplied with the unit. Vital signs parameters include ECG (3 lead /5 lead), SpO2, Respiration, Temperature and Capnography (CO2). It can also display the digital values of HR/PR, SpO2, RR, Non-Invasive Blood Pressure (Systolic, Diastolic and Mean), Temperature, EtCO2, and FiCO2 readings.

In addition, PLANET 55 Model 100 has got Arrhythmia and ST detection from 3L / 5L ECG measurements. The Arrhythmia and ST analysis module is intended for use with Adult & Pediatric patients and is not intended for use with Neonatal Patients.

The user, responsible to interpret the monitored data made available, will be a professional health care provider. The device permits patient monitoring with adjustable alarm limits as well as visible and audible alarm signals. The monitor is not intended for home use.



K090443

P.4/4

**LARSEN & TOUBRO LIMITED**

ELECTRICAL &amp; ELECTRONICS DIVISION - ELECTRONIC PRODUCTS

Mysore Campus, KIADB Industrial Area, Hebbal - Hootagalli, Mysore - 570 018 • Tel: +91 (821) 2405000 • Fax: +91 (821) 2402668

E-Mail:

Ref:

Date: 05.01.2009

Page: 04 of 04

**TECHNOLOGICAL CHARACTERISTICS OF THE DEVICE COMPARED TO  
THE PREDICATE DEVICE**

**Device:** Larsen & Toubro limited make PLANET 55 Model 100 Patient Monitoring System.

**Predicate device:**

- PLANET 55 Patient Monitoring System (Make: L&T Medical Equipments & Systems) / K071472.
- Passport 2 Vital signs monitor with View 12 ECG Analysis Module (Datascope Corp) / K020550.

The parameters available with the Larsen & Toubro Limited make PLANET 55 Model 100 Patient monitoring system are available with the predicate devices "Passport 2 Vital signs monitor with 12 ECG Analysis Module" for Arrhythmia & ST analysis and Larsen & Toubro Limited make PLANET 55 patient monitoring system for other parameters. Comparison of the parameters of PLANET 55 Model 100 to that of the predicate device is given in the "Predicate device comparison table" document.

**Compliance to standards:**

The following international standards are referred.  
IEC 60601-1 Medical Electrical safety  
IEC 60601-1-2 EMC compliance

**Conclusion:**

Based on the Technological characteristics of PLANET 55 Model 100 and its comparison with that of predicate devices Larsen & Toubro Limited believes that their device is substantially equivalent to this predicate Monitors and doesn't pose any additional risk on safety & effectiveness of the device.

(N. Ravindran)

**Head - Design & Development**

21



DEPARTMENT OF HEALTH & HUMAN SERVICES

Public Health Service

Food and Drug Administration  
9200 Corporate Boulevard  
Rockville MD 20850

JUN - 5 2009

Larsen & Toubro Limited  
c/o Ms. Yolanda Smith  
Smith Associates  
1468 Harwell Ave.  
Crofton, MD 21114

Re: K090443

Trade/Device Name: PLANET 55 Model 100

Regulation Number: 21 CFR 870.1025

Regulation Name: Arrhythmia detector and alarm (including ST-segment measurement and alarm)

Regulatory Class: Class II

Product Code: MHX

Dated: May 6, 2009

Received: May 7, 2009

Dear Ms. Smith:

We have reviewed your Section 510(k) premarket notification of intent to market the device referenced above and have determined the device is substantially equivalent (for the indications for use stated in the enclosure) to legally marketed predicate devices marketed in interstate commerce prior to May 28, 1976, the enactment date of the Medical Device Amendments, or to devices that have been reclassified in accordance with the provisions of the Federal Food, Drug, and Cosmetic Act (Act) that do not require approval of a premarket approval application (PMA). You may, therefore, market the device, subject to the general controls provisions of the Act. The general controls provisions of the Act include requirements for annual registration, listing of devices, good manufacturing practice, labeling, and prohibitions against misbranding and adulteration.

If your device is classified (see above) into either class II (Special Controls) or class III (PMA), it may be subject to additional controls. Existing major regulations affecting your device can be found in the Code of Federal Regulations, Title 21, Parts 800 to 898. In addition, FDA may publish further announcements concerning your device in the Federal Register.

Page 2 – Ms. Yolanda Smith


Please be advised that FDA's issuance of a substantial equivalence determination does not mean that FDA has made a determination that your device complies with other requirements of the Act or any Federal statutes and regulations administered by other Federal agencies. You must comply with all the Act's requirements, including, but not limited to: registration and listing (21 CFR Part 807); labeling (21 CFR Part 801); medical device reporting (reporting of medical device-related adverse events) (21 CFR 803); good manufacturing practice requirements as set forth in the quality systems (QS) regulation (21 CFR Part 820); and if applicable, the electronic product radiation control provisions (Sections 531-542 of the Act); 21 CFR 1000-1050.

If you desire specific advice for your device on our labeling regulation (21 CFR Part 801), please go to <http://www.fda.gov/AboutFDA/CentersOffices/CDRH/CDRHOffices/ucm115809.htm> for the Center for Devices and Radiological Health's (CDRH's) Office of Compliance. Also, please note the regulation entitled, "Misbranding by reference to premarket notification" (21 CFR Part 807.97). For questions regarding the reporting of adverse events under the MDR regulation (21 CFR Part 803), please go to <http://www.fda.gov/cdrh/mdr/> for the CDRH's Office of Surveillance and Biometrics/Division of Postmarket Surveillance.

---

You may obtain other general information on your responsibilities under the Act from the Division of Small Manufacturers, International and Consumer Assistance at its toll-free number (800) 638-2041 or (240) 276-3150 or at its Internet address <http://www.fda.gov/cdrh/industry/support/index.html>.

Sincerely yours,



Bram D. Zuckerman, M.D.  
Director  
Division of Cardiovascular Devices  
Office of Device Evaluation  
Center for Devices and  
Radiological Health

Enclosure

### Indications for Use

510(k) Number (if known) \_\_\_\_\_

Device name: **PLANET 55 Model 100**

Indication for use:

The PLANET 55 Model 100 multi-parameter Patient Monitoring system is intended to monitor a single Adult, Pediatric or Neonatal patient's vital signs at the bedside or during intra-hospital transport along with the appropriate accessories mentioned / supplied with the unit. Vital signs parameters include ECG (3 lead /5 lead), SpO2, Respiration, Temperature and Capnography (CO2). It can also display the digital values of HR/PR, SpO2, RR, Non-Invasive Blood Pressure (Systolic, Diastolic and Mean), Temperature, EtCO2, and FiCO2 readings.

In addition, PLANET 55 Model 100 has Arrhythmia and ST detection from 3L / 5L ECG measurements. The Arrhythmia and ST analysis module is intended for use with Adult & Pediatric patients and is not intended for use with Neonatal Patients.

The user, responsible to interpret the monitored data made available, will be a professional health care provider. The device permits patient monitoring with adjustable alarm limits as well as visible and audible alarm signals. The monitor is not intended for home use.

Prescription Use √  
(Part 21 CFR 801 Subpart D)

AND/OR

Over-The-Counter Use \_\_\_\_\_  
(Part 21 CFR 801 Subpart C)

(PLEASE DO NOT WRITE BELOW THIS LINE - CONTINUE ON ANOTHER PAGE IF NEEDED)

Concurrence of CDRH, Office of Device Evaluation (ODE)

*Richard D. B. BZUCKERMAN*  
**(Division Sign-Off)** 6/5/09  
**Division of Cardiovascular Devices**

Page 1 of 1

510(k) Number K090443