

090511

MAQUET

510(k) Summary

[as required by 21 CFR 807.92(c)]

JUN 12 2009

Submitter **MAQUET Cardiopulmonary AG**
Hechinger Strasse 38
72145 Hirrlingen
Germany

Contact Person Frank Moehrke
Phone: 011 49 7478 921 229
Fax: 011 49 7478 921 400

Date Prepared February 24, 2009

Device Trade Name **QUADROX-i Adult microporous membrane
Oxygenator with and without integrated Arterial
Filter with BIOLINE Coating**

Common/Usual Name Oxygenator with integrated heat exchanger and
optional integrated arterial filter

Classification Names Cardiopulmonary bypass oxygenator
(21 CFR 870.4350 – Product Code: DTZ)
Cardiopulmonary bypass heat exchanger
(21 CFR 870.4240 – Product Code: DTR)
Cardiopulmonary bypass arterial line blood filter
(21 CFR 870.4260 – Product Code: DTM)

Legally Marketed Devices - QUADROX-i Adult microporous membrane
Oxygenator with and without integrated
Arterial Filter with Softline Coating (K082117),
- QUART Arterial Filter with BIOLINE Coating
(K082544)

Device Description

The QUADROX-i Adult microporous membrane Oxygenator with and without integrated Arterial Filter with BIOLINE Coating is a blood-gas exchanger with integrated heat exchanger and optionally integrated arterial blood filter.

Indications for Use

The membrane oxygenator QUADROX-i Adult is intended for use in extracorporeal circulation during cardiopulmonary bypass in cardiac surgery. The blood flow rate is defined from 0.5 – 7 l/min. Within the specified flow rate range, the device oxygenates the blood, eliminates carbon dioxide and regulates blood temperature.

MAQUET

The QUADROX-i Adult (HMO 71000) version with integrated arterial filter also filters out air bubbles and particles larger than 40 µm.

The device's utilization period is limited to six hours. The oxygenator is suitable for delivery of the volatile anesthetics isoflurane and sevoflurane. The anesthetic gas is administered through the oxygenator's gas inlet by means of a suitable anesthetic gas vaporizer. Responsibility for deciding whether to use an oxygenator rests solely with the attending physician.

Statement of Technical Comparison

The QUADROX-i Adult microporous membrane Oxygenator with and without integrated Arterial Filter with BIOLINE Coating is identical to the QUADROX-i Adult microporous membrane Oxygenator with and without integrated Arterial Filter with Softline Coating with the only exception that the QUADROX-i Adult with BIOLINE Coating have been coated with BIOLINE Coating instead of Softline Coating. However, the BIOLINE Coating is the same as with the QUART Arterial Filter with BIOLINE Coating. Besides this difference the QUADROX-i Adult microporous membrane Oxygenator with and without integrated Arterial Filter with BIOLINE Coating is the same in design, intended use, method of operation, components, packaging, and fundamental scientific technology as compared to the QUADROX-i Adult microporous membrane Oxygenator with and without integrated Arterial Filter with Softline Coating.

Determination of Substantial Equivalence

Evaluation and testing on safety and effectiveness was executed to demonstrate that the QUADROX-i Adult microporous membrane Oxygenator with and without integrated Arterial Filter with BIOLINE Coating described in this submission is substantially equivalent to the QUADROX-i Adult microporous membrane Oxygenator with and without integrated Arterial Filter with Softline Coating as oxygenator and to the QUART Arterial Filter with BIOLINE Coating regarding the BIOLINE coating.

The following areas have been tested and / or evaluated:

- Integrity
- Performance
- Biocompatibility
- Sterility

Conclusion

The data given demonstrate that the QUADROX-i Adult microporous membrane Oxygenator with and without integrated Arterial Filter with BIOLINE Coating is substantially equivalent to the named predicate devices which hold currently market clearance.



Food and Drug Administration
9200 Corporate Boulevard
Rockville MD 20850

JUN 12 2009

Maquet Cardiopulmonary AG
c/o Mr. Frank Mohrke
Regulatory Affairs Manager
Hechinger Strasse 38
72145 Hirrlingen, Germany

Re: K090511

Quadrox-i Adult Microporous Membrane Oxygenator with and without Integrated Arterial Filter with Bioline Coating
Regulation Number: 21 CFR 870.4350
Regulation Name: Cardiopulmonary bypass oxygenator
Regulatory Class: Class II (two)
Product Code: DTZ
Dated: May 13, 2009
Received: May 15, 2009

Dear Mr. Mohrke:


We have reviewed your Section 510(k) premarket notification of intent to market the device referenced above and have determined the device is substantially equivalent (for the indications for use stated in the enclosure) to legally marketed predicate devices marketed in interstate commerce prior to May 28, 1976, the enactment date of the Medical Device Amendments, or to devices that have been reclassified in accordance with the provisions of the Federal Food, Drug, and Cosmetic Act (Act) that do not require approval of a premarket approval application (PMA). You may, therefore, market the device, subject to the general controls provisions of the Act. The general controls provisions of the Act include requirements for annual registration, listing of devices, good manufacturing practice, labeling, and prohibitions against misbranding and adulteration.

If your device is classified (see above) into either class II (Special Controls) or class III (PMA), it may be subject to such additional controls. Existing major regulations affecting your device can be found in the Code of Federal Regulations, Title 21, Parts 800 to 898. In addition, FDA may publish further announcements concerning your device in the Federal Register.

Please be advised that FDA's issuance of a substantial equivalence determination does not mean that FDA has made a determination that your device complies with other requirements of the Act or any Federal statutes and regulations administered by other Federal agencies. You must comply with all the Act's requirements, including, but not limited to: registration and listing (21 CFR Part 807); labeling (21 CFR Part 801); good manufacturing practice requirements as set forth in the quality systems (QS) regulation (21 CFR Part 820); and if applicable, the electronic product radiation control provisions (Sections 531-542 of the Act); 21 CFR 1000-1050. This letter will allow you to begin marketing your device as described in your Section 510(k) premarket notification. The FDA finding of substantial equivalence of your device to a legally marketed predicate device results in a classification for your device and thus, permits your device to proceed to the market.

If you desire specific advice for your device on our labeling regulation (21 CFR Part 801), please contact the Office of Compliance at (240) 276-0120. Also, please note the regulation entitled, "Misbranding by reference to premarket notification" (21CFR Part 807.97). You may obtain other general information on your responsibilities under the Act from the Division of Small Manufacturers, International and Consumer Assistance at its toll-free number (800) 638-2041 or (240) 276-3150 or at its Internet address <http://www.fda.gov/cdrh/industry/support/index.html>.

Sincerely yours,



 Bram D. Zuckerman, M.D.
Director
Division of Cardiovascular Devices
Office of Device Evaluation
Center for Devices and
Radiological Health

Enclosure

Indications for Use

510(k) Number (if known): K090511

Device Name:

**QUADROX-i Adult microporous membrane Oxygenator
with and without integrated Arterial Filter with BIOLINE Coating**

Indications for Use:

The membrane oxygenator QUADROX-i Adult is intended for use in extracorporeal circulation during cardiopulmonary bypass in cardiac surgery. The blood flow rate is defined from 0.5 – 7 l/min. Within the specified flow rate range, the device oxygenates the blood, eliminates carbon dioxide and regulates blood temperature.

The QUADROX-i Adult (HMO 71000) version with integrated arterial filter also filters out air bubbles and particles larger than 40 µm.

The device's utilization period is limited to six hours. The oxygenator is suitable for delivery of the volatile anesthetics isoflurane and sevoflurane. The anesthetic gas is administered through the oxygenator's gas inlet by means of a suitable anesthetic gas vaporizer. Responsibility for deciding whether to use an oxygenator rests solely with the attending physician.

Prescription Use X AND/OR Over-The-Counter Use _____
(Part 21 CFR 801 Subpart D) (21 CFR 801 Subpart C)

(PLEASE DO NOT WRITE BELOW THIS LINE-CONTINUE ON ANOTHER PAGE IF NEEDED)

Concurrence of CDRH, Office of Device Evaluation (ODE)

Page 1 of 1

Danna R. Volmer
(Division Sign-Off)
Division of Cardiovascular Devices

510(k) Number K090511