510(K) SUMMARY

This summary of 510(k) safety and effectiveness information is being submitted in accordance with the requirements of SMDA 1990 and 21 CFR §807.92(c).

The assigned 510(k) number is: K090912.

1. Submitter:

Shenzhen Mindray Bio-medical Electronics Co., LTD Mindray Building, Keji 12th Road South, Hi-tech Industrial Park, Nanshan, Shenzhen, 518057, P. R. China

Tel: +86 755 2658 2888 Fax: +86 755 2658 2680

Contact Person:

Tan Chuanbin Shenzhen Mindray Bio-medical Electronics Co., LTD Mindray Building, Keji 12th Road South, Hi-tech Industrial Park, Nanshan, Shenzhen, 518057, P. R. China

Date Prepared: March 6, 2009

2. <u>Device Name</u>: DP-6900 Digital Ultrasonic Diagnostic Imaging System

Classification

Regulatory Class: II
Review Category: Tier II

21 CFR 892.1560 Ultrasonic Pulsed Echo Imaging System (90-IYO)

21 CFR 892.1570 Diagnostic Ultrasound Transducer (90-ITX)

3. Marketed Device:

DP-6900 Digital Ultrasonic Diagnostic Imaging System is substantially equivalent to the following devices: Mindray DC-6 (K#072164), Mindray DP-9900 (K#070526), Mindray DP-6600 (K#060949) and GE Logiq 9 (K061129).

4. Device Description:

The DP-6900 Digital Ultrasonic Diagnostic Imaging System is a general purpose,

portable, software controlled, ultrasound diagnostic system. Its function is to acquire and display ultrasound images in B-Mode, M-Mode or the combined mode (i.e. B/M Mode). This system is a Track 1 device that employs an array of probes that include linear array and convex array with a frequency range of approximately 2.0 MHz to 10.0 MHz.

5. Intended Use:


The device is intended for use by a qualified physician for ultrasound evaluation of abdominal, cardiac, small parts (breast, testes, thyroid, etc.), urology, peripheral vascular, fetal, transrectal, transvaginal, intraoperative, pediatric, neonatal cephalic, musculoskeletal (general and superficial).

6. Safety Considerations:

The DP-6900 Digital Ultrasonic Diagnostic Imaging System had been tested as Track 1 Device per the FDA Guidance document "Information for Manufacturers Seeking Marketing Clearance of Diagnostic Ultrasound Systems and Transducers" issued in September 2008. The acoustic output is measured and calculated per NEMA UD 2: 2004 Acoustic Output Measurement Standard for Diagnostic Ultrasound Equipment. The device conforms to applicable medical device safety standards, such as IEC 60601-1, IEC 60601-1-1, IEC 60601-1-2, IEC 60601-2-37, IEC 60601-1-4 and ISO 10993-1.

Conclusion:

The conclusions drawn from testing of the DP-6900 Digital Ultrasonic Diagnostic Imaging System demonstrate that the device is as safe and effective as the legally marketed predicate devices.





Food and Drug Administration 9200 Corporate Boulevard Rockville MD 20850

APR 1 5 2009

Shenzhen Mindray Bio-Medical Electronics Co., Ltd. % Mr. Robert Mosenkis President CITECH 5200 Butler Pike Plymouth Meeting, PA 19462-1298

Re: K090912

Trade/Device Name: DP-6900 Digital Ultrasonic Diagnostic Imaging System

Regulation Number: 21 CFR 892.1560

Regulation Name: Ultrasonic pulsed echo imaging system

Regulatory Class: II

Product Code: IYO and ITX Dated: March 31, 2009 Received: April 1, 2009

Dear Mr. Mosenkis:

We have reviewed your Section 510(k) premarket notification of intent to market the device referenced above and we have determined the device is substantially equivalent (for the indications for use stated in the enclosure) to legally marketed predicate devices marketed in interstate commerce prior to May 28, 1976, the enactment date of the Medical Device Amendments, or to devices that have been reclassified in accordance with the provisions of the Federal Food, Drug, and Cosmetic Act (Act). You may, therefore, market the device, subject to the general controls provisions of the Act. The general controls provisions of the Act include requirements for annual registration, listing of devices, good manufacturing practice, labeling, and prohibitions against misbranding and adulteration.

This determination of substantial equivalence applies to the following transducers intended for use with the DP-6900 Digital Ultrasonic Diagnostic Imaging System, as described in your premarket notification:

Transducer Model Number

35C20EA 35C50EA 65EC10EA 65C15EA 65EL60EA 75L38EA 75LT38EA 75L53EA 75L60EA

If your device is classified (see above) into either class II (Special Controls) or class III (PMA), it may be subject to such additional controls. Existing major regulations affecting your device can be found in the Code of Federal Regulations, Title 21, Parts 800 to 898. In addition, FDA may publish further announcements concerning your device in the <u>Federal Register</u>.

Please be advised that FDA's issuance of a substantial equivalence determination does not mean that FDA has made a determination that your device complies with other requirements of the Act or any Federal statutes and regulations administered by other Federal agencies. You must comply with all the Act's requirements, including, but not limited to: registration and listing (21 CFR Part 807); labeling (21 CFR Part 801); good manufacturing practice requirements as set forth in the quality systems (QS) regulation (21 CFR Part 820); and if applicable, the electronic product radiation control provisions (Sections 531-542 of the Act); 21 CFR 1000-1050.

This letter will allow you to begin marketing your device as described in your premarket notification. The FDA finding of substantial equivalence of your device to a legally marketed predicate device results in a classification for your device and thus permits your device to proceed to market.

If you desire specific advice for your device on our labeling regulation (21 CFR Part 801), please contact the Office of Compliance at (240) 276-0120. Also, please note the regulation entitled, "Misbranding by reference to premarket notification" (21 CFR Part 807.97). You may obtain other general information on your responsibilities under the Act from the Division of Small Manufacturers, International and Consumer Assistance at its toll-free number (800) 638-2041 or (240) 276-3150 or at its Internet address http://www.fda.gov/cdrh/industry/support/index.html

If you have any questions regarding the content of this letter, please contact Paul Hardy at (240) 276-3666.

Sincerely yours,

Janine M. Morris

Acting Director, Division of Reproductive, Abdominal, and Radiological Devices Office of Device Evaluation

Center for Devices and Radiological Health

Enclosure(s)

System	×	Transducer	
Model:	DP-6900	_	
510(k) Number(s)	1090912	_ _	

Clin	ical Application								
General	Specific	В	М	PWD	CWD	Color Doppler	Amplitude Doppler	Combined (specify)	Other (specify)
Ophthalmic	Ophthalmic								
	Fetal	N	N					N	Note 1
	Abdominal	z	N					N	Note 1
	Intra-operative (Specify)*	N	N					N	
	Intra-operative (Neuro)								
	Laparoscopic								
	Pediatric	N	N					N	Note 1
	Small Organ(Specify)**	N	N					N	
•	Neonatal Cephalic	N	N					N	
Fetal Imaging & Other	Adult Cephalic	N	N	•				N	
	Trans-rectal	N	N			Ü		N	
	Trans-vaginal	N	N					N	····
	Trans-urethral								-
	Trans-esoph.(non-Card.)								
	Musculo-skeletal (Conventional)	N	N		-			N	Note 1
	Musculo-skeletal (Superficial)	N	N					N	
	Intravascular								
	Other (specify)								
	Cardiac Adult	N	Z					N	
	Cardiac Pediatric	N	N					N	
Cardiac	Intravascular(Cardiac)								
Cardiac	Trans-esoph.(Cardiac)								
	Intra-cardiac						<u>_</u>		·
	Other (specify)								
Peripheral vessel	Peripheral vessel	N	N.				<u> </u>	N	Note 1
oribuciai resset	Other(Specify)								

N=new	indication; P=	previously cle	ared by FDA;	E=added ur	ider Appendix E	•	

Additional comments: Combined modes: B+M.	
*Intraoperative includes abdominal, thoracic, and vascular etc.	
**Small organ-breast, thyroid, testes, etc.	,

Note 1: Tissue Harmonic Imaging. The feature does not use contrast agents.

(PLEASE DO NOT WRITE BELOW THIS LINE-CONTINUE ON ANOTHER PAGE IF NEEDED)

Concurrence of CDRH, Office of Device Evaluation(ODE)

Prescription USE (Per 21 CFR 801.109)

(Division Sign-Off)

Division of Reproductive, Abdominal and

Radiological Devices

510(k) Number _

090912

System		Transducer	×	
Model:	35C20EA			
510(k) Number(s)	K0909/2			

Clini	cal Application								
General	Specific	В	М	PWD	CWD	Color Doppler	Amplitude Doppler	Combined (specify)	Other (specify)
Ophthalmic	Ophthalmic								
,	Fetal						•		
	Abdominal	И	N					N	
	Intra-operative (Specify)*								
	Intra-operative (Neuro)								
	Laparoscopic		I						
	Pediatric	N	N			,		. И	
	Small Organ(Specify)**								
•	Neonatal Cephalic								
Fetal Imaging & Other	Adult Cephalic								
Tomi imaging & Outer	Trans-rectal								
	Trans-vaginal								
	Trans-urethral								
	Trans-esoph.(non-Card.)								
	Musculo-skeletal (Conventional)	N	N		i			И	
	Musculo-skeletal (Superficial)								
	Intravascular								
	Other (specify)								<u> </u>
	Cardiac Adult	N	N					N	
	Cardiac Pediatric	N	N					N	
Cardiac	Intravascular(Cardiac)								
Jaruide	Trans-esoph.(Cardiac)					-			
	Intra-cardiac								
	Other (specify)								
Peripheral vessel	Peripheral vessel	N	N					N	
renpheral vessei	Other(Specify)								

N=new indication; P=previously cleared by FDA; E=added under Appendix E

Additional comments: Combined modes: B+M.

Note 1: Tissue Harmonic Imaging. The feature does not use contrast agents.

(PLEASE DO NOT WRITE BELOW THIS LINE-CONTINUE ON ANOTHER PAGE IF NEEDED)

Concurrence of CDRH, Office of Device Evaluation(ODE)

Prescription USE (Per 21 CFR 801.109)

(Division Sign-Off)

Division of Reproductive, Abdominal and

Radiological Devices

510(k) Number <u>K090912</u>

^{*}Intraoperative includes abdominal, thoracic, and vascular etc.

^{**}Small organ-breast, thyroid, testes, etc.

System		Transducer	×
Model:	35C50EA		
510(k) Number(s)	K090912		

Clini	cal Application								
General	Specific	В	М	PWD	CWD	Color Doppler	Amplitude Doppler	Combined (specify)	Other (specify)
Ophthalmic	Ophthalmic								
. •	Fetal	N	N					N	Note 1
	Abdominal	N	N					N	Note 1
	Intra-operative (Specify)*								
	Intra-operative (Neuro)								
	Laparoscopic								
	Pediatric	N	N					N	Note !
	Small Organ(Specify)**								
Fetal Imaging & Other	Neonatal Cephalic								
	Adult Cephalic								
	Trans-rectal								
	Trans-vaginal								•
	Trans-urethral		I						•
•	Trans-esoph.(non-Card.)								
	Musculo-skeletal (Conventional)	N	N					N	Note 1
	Musculo-skeletal (Superficial)								
	Intravascular								• •
	Other (specify)								
-	Cardiac Adult								
	Cardiac Pediatric								
Cardiac	Intravascular(Cardiac)								
Cardiac	Trans-esoph.(Cardiac)								
	Intra-cardiac								
	Other (specify)								
Peripheral vessel	Peripheral vessel	N	N					N	Note I
culturian sesser	Other(Specify)								•

N=new indication; P=previously	cleared by FDA; E=added under Appendix E

Additional	comment	s:Combined	modes: B+M.	

Note 1: Tissue Harmonic Imaging. The feature does not use contrast agents.

(PLEASE DO NOT WRITE BELOW THIS LINE-CONTINUE ON ANOTHER PAGE IF NEEDED)

Concurrence of CDRH, Office of Device Evaluation (ODE)

Prescription USE (Per 21 CFR 801.109)

(Division Sign-Off)

Division of Reproductive, Abdominal and

Radiological Devices

510(k) Number <u>K0909/Z</u>

^{*}Intraoperative includes abdominal, thoracic, and vascular etc.

^{**}Small organ-breast, thyroid, testes, etc.

System				Transd	ucer		×		
Model:	6SEC10EA								
510(k) Number(s)	K090912	•							
Člini	cal Application								
General	Specific	В	м	PWD	CWD	Color Doppler	Amplitude Doppler	Combined (specify)	Other (specify)
Ophthalmic	Ophthalmic								
	Fetal	N	N					N	
	Abdominal								
	Intra-operative (Specify)*								
	Intra-operative (Neuro)								
Fetal Imaging & Other	Laparoscopic								
	Pediatric								
	Small Organ(Specify)**								
	Neonatal Cephalic	N	N					N	
	Adult Cephalic								
	Trans-rectal	N	N					N	
	Trans-vaginal	N	N					И	
	Trans-urethral								
	Trans-esoph.(non-Card.)								
	Musculo-skeletal (Conventional)								
	Musculo-skeletal (Superficial)								
	Intravascular								
	Other (specify)								
	Cardiac Adult								
	Cardiac Pediatric								
Cardiaç	Intravascular(Cardiac)								
Cardiac	Trans-esoph.(Cardiac)								
	Intra-cardiac				l				
	Other (specify)								
Peripheral vessel	Peripheral vessel								
renpheral vesser	Other(Specify)								
N=new indication; P=pr	eviously cleared by FDA; E=added	i unde	т Арр	endix E					
Additional comments:C	ombined modes: B+M.								
*Intraoperati	ve includes abdominal, thoracic, ar	nd vas	cular	etc.					
**Small orga	n-breast thyroid testes etc.								

Note 1: Tissue Harmonic Imaging. The feature does not use contrast agents.

(PLEASE DO NOT WRITE BELOW THIS LINE-CONTINUE ON ANOTHER PAGE IF NEEDED)

Concurrence of CDRH, Office of Device Evaluation(ODE)

Prescription USE (Per 21 CFR 801.109)

(Division Sign-Off)

Division of Reproductive, Abdominal and

Radiological Devices

510(k) Number <u>K0909/</u>

System				Transd	ucer		×		
Model: 510(k) Number(s)	KO90912								
C	linical Application								
	0.00	-	Г.,	num	CVID	Color	Amplitude	Combined	

Clinical Application									
General	Specific	В	М	PWD	CWD	Color Doppler	Amplitude Doppler	Combined (specify)	Other (specify)
Ophthalmic	Ophthalmic								
	Fetal								
	Abdominal	N	N			-		N	
	Intra-operative (Specify)*								·
	Intra-operative (Neuro)								
	Laparoscopic								
	Pediatric	N	N					N	
	Small Organ(Specify)**								
	Neonatal Cephalic	N	N					N	
Fetal Imaging & Other	Adult Cephalic	N	N					И	
Cut maging & Outer	Trans-rectal								
	Trans-vaginal								
	Trans-urethral								
	Trans-esoph.(non-Card.)								
	Musculo-skeletal (Conventional)	N	N					N	
	Musculo-skeletal (Superficial)						_		
	Intravascular								
	Other (specify)		1						
	Cardiac Adult								
	Cardiac Pediatric	N	N					N	
Cardiac	Intravascular(Cardiac)								
Cardiac	Trans-esoph.(Cardiac)								
	Intra-cardiac								
	Other (specify)								
Peripheral vessel	Peripheral vessel	Ŋ	N					N	
r cribiiciai acasei	Other(Specify)								•

N=new indication; P=previously cleared by FDA; E=added under Appendix E

Additional comments: Combined modes: B+M.

*Intraoperative includes abdominal, thoracic, and vascular etc.

**Small organ-breast, thyroid, testes, etc.

Note 1: Tissue Harmonic Imaging. The feature does not use contrast agents.

(PLEASE DO NOT WRITE BELOW THIS LINE-CONTINUE ON ANOTHER PAGE IF NEEDED)

Concurrence of CDRH, Office of Device Evaluation(ODE)

Prescription USE (Per 21 CFR 801.109)

(Division Sign-Off)

Division of Reproductive, Abdominal and

Radiological Devices

510(k) Number <u>KO 90 9/2</u>

System	4000 4000			t ranso	nce.		<u> </u>		
Model:	65EL60EA								
510(k) Number(s)	K090912								
Clini	cal Application					•			
General	Specific	В	М	PWD	CWD	Color Doppler	Amplitude Doppler	Combined (specify)	Other (specify)
Ophthalmic	Ophthalmic								
	Fetal								
•	Abdominal			,					
	Intra-operative (Specify)*								
	Intra-operative (Neuro)								
	Laparoscopic								
	Pediatric								
	Small Organ(Specify)**								
	Neonatal Cephalic								
Fetal Imaging & Other	Adult Cephalic						1		
retat imaging & Other	Trans-rectal	N	N					N	
	Trans-vaginal								
	Trans-urethral						<u> </u>		
	Trans-esoph.(non-Card.)								
	Musculo-skeletal (Conventional)					·			_
	Musculo-skeletal (Superficial)								
	Intravascular								
	Other (specify)						<u> </u>		
	Cardiac Adult								
	Cardiac Pediatric								
Cardiac	Intravascular(Cardiac)								
Cardiac	Trans-esoph.(Cardiac)]					
	Intra-cardiac								•
	Other (specify)							,	
Peripheral vessel	Peripheral vessel								
Cripacial vesser	Other(Specify)								
N=new indication; P=pr	eviously cleared by FDA; E=added	unde	т Арр	endix E			· · · · · · · · · · · · · · · · · · ·		
Additional comments:C	ombined modes: B+M.								
*Intraoperati	ve includes abdominal, thoracic, as	nd vas	cular	etc.					
**Small orga	in-breast, thyroid, testes, etc.								
	ue Harmonic Imaging. The feature	does r	iot use	contras	t agents.				•

(PLEASE DO NOT WRITE BELOW THIS LINE-CONTINUE ON ANOTHER PAGE IF NEEDED) Concurrence of CDRH, Office of Device Evaluation(ODE)

Prescription USE (Per 21 CFR 801.109) (Division Sign-Off) Division of Reproductive, Abdominal and Radiological Devices 510(k) Number _

System				Transdi	icer		×		
Model:	75L38EA								
510(k) Number(s)	K090912								
Clini	cal Application				,	· ·			
General	Specific	В	М	PWD	CWD	Color Doppler	Amplitude Doppler	Combined (specify)	Other (specify)
Ophthalmic	Ophthalmic			·	ļ				
	Fetal					<u> </u>	<u> </u>		
	Abdominal	N	N	<u> </u>				N	
	Intra-operative (Specify)*								
	Intra-operative (Neuro)								
	Laparoscopic			<u> </u>]			
	Pediatric	N	N					N	
	Small Organ(Specify)**	N	N					N	
	Neonatal Cephalic	N	N				<u> </u>	N	
Carl Impaire & Other	Adult Cephalic]	ļ	
Fetal Imaging & Other	Trans-rectal							<u> </u>	
	Trans-vaginal			I	}			<u> </u>	
	Trans-urethral							<u> </u>	
	Trans-esoph (non-Card.)								
	Musculo-skeletal (Conventional)	N	И				<u> </u>	N	
	Musculo-skeletal (Superficial)	N	N	Π			<u> </u>	N ·	
	Intravascular	1				T		<u> </u>	<u></u>
	Other (specify)							<u> </u>	
****	Cardiac Adult			Ţ	Ι΄				
	Cardiac Pediatric	1	abla	Ī				<u> </u>	
	Intravascular(Cardiac)				j				<u> </u>
Cardiac	Trans-esoph.(Cardiac)							<u> </u>	<u> </u>
	Intra-cardiac	T							
1	Other (specify)		T						
	Peripheral vessel	N	N]	N	
Peripheral vessel	Other(Specify)							<u>.l</u>	
N=new indication; P=p	reviously cleared by FDA; E=adde	d und	er Ap	pendix l	Ε				_
	Combined modes: B+M.								_
	tive includes abdominal, thoracic, a	ınd va	scula	etc.		,			_
	gan-breast, thyroid, testes, etc.								- -
Note 1: Tiss	sue Harmonic Imaging. The feature	does	not u	se contr	ast agent	s.			

(PLEASE DO NOT WRITE BELOW THIS LINE-CONTINUE ON ANOTHER PAGE IF NEEDED)

Concurrence of CDRH, Office of Device Evaluation(ODE)

Prescription USE (Per 21 CFR 801.109)
(Division Sign-Off)

Division of Reproductive, Abdominal and

Radiological Devices

510(k) Number <u>K09091Z</u>

System		Transducer	×	
Model:	75LT38EA			
510(k) Number(s)	K0909 12			

Clini	cal Application								
General	Specific	В	М	PWD	CWD	Color Doppler	Amplitude Doppler	Combined (specify)	Other (specify)
Ophthalmic	Ophthalmic								
	Fetal .								
	Abdomina!	N	N					N	
	Intra-operative (Specify)*	N	N					N	
	Intra-operative (Neuro)								
	Laparoscopic								
	Pediatric	N	N					N	
•	Small Organ(Specify)**	N	N					N	
	Neonatal Cephalic	N	N					Ŋ	
Fetal Imaging & Other	Adult Cephalic					L			
	Trans-rectal			Ĺ <u></u> .	<u> </u>	<u> </u>			
	Trans-vaginal						ļ		
	Trans-urethral								
	Trans-esoph.(non-Card.)				L			<u> </u>	
	Musculo-skeletal (Conventional)	N	N					N	
	Musculo-skeletal (Superficial)	Z	N					N	
	Intravascular				<u> </u>		<u> </u>	<u></u>	
	Other (specify)			}			<u> </u>		
-	Cardiac Adult			L		<u> </u>			<u> </u>
	Cardiac Pediatric					<u> </u>	<u> </u>		
Cardiac	Intravascular(Cardiac)				<u> </u>	<u> </u>	<u> </u>	<u> </u>	
Cardiac	Trans-esoph (Cardiac)		<u> </u>			<u> </u>		ļ	<u> </u>
	Intra-cardiac								
·	Other (specify)					<u> </u>	<u> </u>		
Parinhard vessel	Peripheral vessel	N	N		<u> </u>	<u> </u>	<u> </u>	N	
Peripheral vessel	Other(Specify)			1			1	1	! .

N=new indication; P=previously cleared by FDA; E=added under Appendix	E		
Additional comments:Combined modes: B+M.			·
*Intraoperative includes abdominal, thoracic, and vascular etc.		· · · · · · · · · · · · · · · · · · ·	
**Small organ-breast, thyroid, testes, etc.			

Note 1: Tissue Harmonic Imaging. The feature does not use contrast agents.

(PLEASE DO NOT WRITE BELOW THIS LINE-CONTINUE ON ANOTHER PAGE IF NEEDED)

Concurrence of CDRH, Office of Device Evaluation(ODE)

(Division Sign-Off)
Division of Reproductive, Abdominal and Radiological Devices

510(k) Number .

System				Transdu	ıcer		×			
Model: 510(k) Number(s)	75L53EA K090912									
Clini	cal Application							7		
General	Specific	В	М	PWD	CWD	Color Doppler	Amplitude Doppler	Combined (specify)	Other (specify)	
Ophthalmic	Ophthalmic									
	Fetal						ļ		<u> </u>	
	Abdominal	N	N				<u> </u>	N		
	Intra-operative (Specify)*						<u> </u>			
	Intra-operative (Neuro)			L	<u> </u>		\	<u></u>		
	Laparoscopic		<u> </u>							
	Pediatric	N	N					N		
	Small Organ(Specify)**	Ŋ	N	<u> </u>		<u> </u>	<u> </u>	N		
	Neonatal Cephalic	N	N				<u> </u>	N ·	ļ	
Fetal Imaging & Other	Adult Cephalic				<u> </u>	<u> </u>		<u> </u>		
retal thaging of Other	Trans-rectal				<u> </u>		ļ	<u> </u>		
	Trans-vaginal				<u> </u>				<u> </u>	
	Trans-urethral		<u> </u>			<u> </u>	<u> </u>			
	Trans-esoph.(non-Card.)			<u> </u>						
	Musculo-skeletal (Conventional)	N	N					N		
	Musculo-skeletal (Superficial)	N	N					N		
1	Intravascular					<u> </u>				
	Other (specify)		<u>L</u>		<u> </u>		<u> </u>	<u> </u>		
	Cardiac Adult			<u> </u>			<u> </u>			
	Cardiac Pediatric				<u>L.</u>	<u> </u>				
	Intravascular(Cardiac)				<u> </u>			<u> </u>		
Cardiac	Trans-esoph.(Cardiac)					<u> </u>		ļ		
]	Intra-cardiac						<u> </u>			
	Other (specify)									
	Peripheral vessel	N	N					N		
Peripheral vessel	Other(Specify)			Ī.,		<u> </u>	<u> </u>		<u> </u>	
N=new indication; P=p	oreviously cleared by FDA; E=adde	d und	er Ap	pendix I	2				-	
- '	Combined modes: B+M.								_	
	tive includes abdominal thoracic a	nd va	coular	etc						

Note 1: Tissue Harmonic Imaging. The feature does not use contrast agents.

**Small organ-breast, thyroid, testes, etc.

(PLEASE DO NOT WRITE BELOW THIS LINE-CONTINUE ON ANOTHER PAGE IF NEEDED)

Concurrence of CDRH, Office of Device Evaluation(ODE)

(Division Sign-Off)
Division of Reproductive, Abdominal and Radiological Devices
510(k) Number

System				Transu	1001				
Model:	75L60EA								
510(k) Number(s)	K0909 12								
Clini	cal Application								
General	Specific	В	М	PWD	CWD	Color Doppler	Amplitude Doppler	Combined (specify)	Other (specify)
Ophthalmic	Ophthalmic								
* "-	Fetal							<u> </u>	
	Abdominal	N	N			·		N	
	Intra-operative (Specify)*								
	Intra-operative (Neuro)								·
	Laparoscopic							ļ	 -
	Pediatric	N	N		<u> </u>			N	
	Small Organ(Specify)**	N	N				<u> </u>	N	
	Neonatal Cephalic	N	N				<u> </u>	N	
Fetal Imaging & Other	Adult Cephalic			<u> </u>		<u> </u>			
i viai iiiiagiiig ce o iiivi	Trans-rectal	·		1	<u> </u>		<u> </u>		
	Trans-vaginal			<u> </u>		<u></u>	ļ	<u> </u>	
	Trans-urethral								
	Trans-esoph (non-Card.)								
	Musculo-skeletal (Conventional)	N	N					N	
	Musculo-skeletal (Superficial)	N	N					N	
	Intravascular								
	Other (specify)								
	Cardiac Adult		1		1			1	

N=new indication; P=previously cleared by FDA; E=added under Appendix E

Additional comments: Combined modes: B+M.

*Intraoperative includes abdominal, thoracic, and vascular etc.

**Small organ-breast, thyroid, testes, etc.

Other(Specify)

Cardiac Pediatric
Intravascular(Cardiac)

Trans-esoph.(Cardiac)
Intra-cardiac
Other (specify)
Peripheral vessel

Cardiac

Peripheral vessel

Note 1: Tissue Harmonic Imaging. The feature does not use contrast agents.

(PLEASE DO NOT WRITE BELOW THIS LINE-CONTINUE ON ANOTHER PAGE IF NEEDED)

Concurrence of CDRH, Office of Device Evaluation(ODE)

(Division of Reproductive, Abdominal and

Radiological Devices 510(k) Number