

K091252

20

9 SMDA Summary of Safety and Effectiveness – “510 (k) Summary”

JUL 22 2009

A. Submitter Information

SATELEC
17, Avenue Gustave Eiffel
BP 30216
33708 Merignac Cedex
FRANCE

Telephone: 011 33 556 34 0607
Fax: 011 33 556 34 9292

Contact Person: Rick Rosati
SATELEC
c/o ACTEON, Inc.
124 Gaither Drive, Suite 140
Mt. Laurel, NJ 08054
Tel: 800 289-6367 Ext. 39
Fax: 856 222-4726
Email: rick.rosati@us.acteongroup.com

Date Prepared: 04 / 30 / 2009 **REVISED 7/16/09**

B. Device Identification

Common Usual Name: *“Bone Cutting Instrument and Accessories”*

Proprietary Name: Implant Center 2

C. Identification of the Predicate Device

<u>Device</u>	<u>Applicant</u>	<u>510(k) No.</u>	<u>Date Cleared</u>
Implant Center	Satelec	K072030	11 / 19 / 2007

The Satelec Implant Center 2 is substantially equivalent to the predicate device by Satelec, the Implant Center (K072030) previously cleared by the FDA and currently marketed.

D. Device Description

The Satelec Implant Center 2 is a dental operative unit that supplies utilities to and serves as a base for dental tools and accessories for use by qualified dental practitioners.

E. Substantial Equivalence

The Implant Center 2 and the predicate device, Implant Center (K072030) are both dental operative units that supply utilities to and serve as a base for dental tools and accessories for use by qualified dental practitioners. Differences that exist between the devices relating to technical specification, performances and intended use are minor and do not affect the safety and effectiveness of the Implant Center 2.



Food and Drug Administration
9200 Corporate Boulevard
Rockville MD 20850

SATELEC
C/O Mr. Rick Rosati
Quality Manager
ACTEON, Incorporated
124 Gaither Drive, Suite 140
Mount Laurel, New Jersey 08054

JUL 22 2009

Re: K091252
Trade/Device Name: IMPLANT CENTER 2
Regulation Number: 21 CFR 872.4120
Regulation Name: Bone Cutting Material and Accessories
Regulatory Class: II
Product Code: DZI
Dated: April 29, 2009
Received: April 29, 2009

Dear Mr. Rosati:

We have reviewed your Section 510(k) premarket notification of intent to market the device referenced above and have determined the device is substantially equivalent (for the indications for use stated in the enclosure) to legally marketed predicate devices marketed in interstate commerce prior to May 28, 1976, the enactment date of the Medical Device Amendments, or to devices that have been reclassified in accordance with the provisions of the Federal Food, Drug, and Cosmetic Act (Act) that do not require approval of a premarket approval application (PMA). You may, therefore, market the device, subject to the general controls provisions of the Act. The general controls provisions of the Act include requirements for annual registration, listing of devices, good manufacturing practice, labeling, and prohibitions against misbranding and adulteration.

If your device is classified (see above) into either class II (Special Controls) or class III (PMA), it may be subject to additional controls. Existing major regulations affecting your device can be found in the Code of Federal Regulations, Title 21, Parts 800 to 898. In addition, FDA may publish further announcements concerning your device in the Federal Register.

Please be advised that FDA's issuance of a substantial equivalence determination does not mean that FDA has made a determination that your device complies with other requirements of the Act or any Federal statutes and regulations administered by other Federal agencies. You must comply with all the Act's requirements, including, but not limited to: registration and listing (21 CFR Part 807); labeling (21 CFR Part 801); medical device reporting (reporting of medical device-related adverse events) (21 CFR 803); good manufacturing practice requirements as set forth in the quality systems (QS) regulation (21 CFR Part 820); and if applicable, the electronic product radiation control provisions (Sections 531-542 of the Act); 21 CFR 1000-1050.

If you desire specific advice for your device on our labeling regulation (21 CFR Part 801), please go to

<http://www.fda.gov/AboutFDA/CentersOffices/CDRH/CDRHOffices/ucml15809.htm> for the Center for Devices and Radiological Health's (CDRH's) Office of Compliance. Also, please note the regulation entitled, "Misbranding by reference to premarket notification" (21CFR Part 807.97). For questions regarding the reporting of adverse events under the MDR regulation (21 CFR Part 803), please go to <http://www.fda.gov/cdrh/mdr/> for the CDRH's Office of Surveillance and Biometrics/Division of Postmarket Surveillance.

You may obtain other general information on your responsibilities under the Act from the Division of Small Manufacturers, International and Consumer Assistance at its toll-free number (800) 638-2041 or (240) 276-3150 or at its Internet address <http://www.fda.gov/cdrh/industry/support/index.html>.

Sincerely yours,



Susan Runner, D.D.S., M.A.

Acting Director

Division of Anesthesiology, General Hospital,

Infection Control and Dental Devices

Office of Device Evaluation

Center for Devices and

Radiological Health

Enclosure

Indications for Use

510(k) Number: _____

Device Name: **IMPLANT CENTER 2**

Indications for Use:

"The intended use of the Satelec Implant Center 2 is to supply utilities to and serve as a base for dental tools and accessories for use by qualified dental practitioners.."


Prescription Use X
(Part 21 CFR 801 Subpart D)

AND/OR

Over-The-Counter Use _____
(21 CFR 801 Subpart C)

(PLEASE DO NOT WRITE BELOW THIS LINE – CONTINUE ON ANOTHER PAGE
IF NEEDED)

Concurrence of CDRH, Office of Device Evaluation (ODE)



(Division Sign-Off)
Division of Anesthesiology, General Hospital
Infection Control, Dental Devices

510(k) Number: K091252