



Food and Drug Administration
10903 New Hampshire Avenue
Document Control Room –WO66-0609
Silver Spring, MD 20993-0002

Alden Optical Laboratories
C/O Charles H. Creighton
CEO
13295 Broadway
Alden, NY 14004

JUN 25 2010

Re: K091327

Trade/Device Name: Alden HP 54 (hioxifilcon D) Spherical, Toric, Multifocal and Toric Multifocal Contact Lenses
Regulation Number: 21 CFR 886.5925
Regulation Name: Soft (hydrophilic) contact lens
Regulatory Class: Class II
Product Code: LPL
Dated: March 19, 2010
Received: March 23, 2010

Dear Mr. Creighton:

We have reviewed your Section 510(k) premarket notification of intent to market the device referenced above and have determined the device is substantially equivalent (for the indications for use stated in the enclosure) to legally marketed predicate devices marketed in interstate commerce prior to May 28, 1976, the enactment date of the Medical Device Amendments, or to devices that have been reclassified in accordance with the provisions of the Federal Food, Drug, and Cosmetic Act (Act) that do not require approval of a premarket approval application (PMA). You may, therefore, market the device, subject to the general controls provisions of the Act. The general controls provisions of the Act include requirements for annual registration, listing of devices, good manufacturing practice, labeling, and prohibitions against misbranding and adulteration. Please note: CDRH does not evaluate information related to contract liability warranties. We remind you, however, that device labeling must be truthful and not misleading.

If your device is classified (see above) into either class II (Special Controls) or class III (PMA), it may be subject to additional controls. Existing major regulations affecting your device can be found in the Code of Federal Regulations, Title 21, Parts 800 to 898. In addition, FDA may publish further announcements concerning your device in the Federal Register.

Please be advised that FDA's issuance of a substantial equivalence determination does not mean that FDA has made a determination that your device complies with other requirements of the Act or any Federal statutes and regulations administered by other Federal agencies. You must


Page 2 – Mr. Charles H. Creighton

comply with all the Act's requirements, including, but not limited to: registration and listing (21 CFR Part 807); labeling (21 CFR Part 801); medical device reporting (reporting of medical device-related adverse events) (21 CFR 803); good manufacturing practice requirements as set forth in the quality systems (QS) regulation (21 CFR Part 820); and if applicable, the electronic product radiation control provisions (Sections 531-542 of the Act); 21 CFR 1000-1050.

If you desire specific advice for your device on our labeling regulation (21 CFR Part 801), please go to <http://www.fda.gov/AboutFDA/CentersOffices/CDRH/CDRHOffices/ucm115809.htm> for the Center for Devices and Radiological Health's (CDRH's) Office of Compliance. Also, please note the regulation entitled, "Misbranding by reference to premarket notification" (21CFR Part 807.97). For questions regarding the reporting of adverse events under the MDR regulation (21 CFR Part 803), please go to <http://www.fda.gov/MedicalDevices/Safety/ReportaProblem/default.htm> for the CDRH's Office of Surveillance and Biometrics/Division of Postmarket Surveillance.

You may obtain other general information on your responsibilities under the Act from the Division of Small Manufacturers, International and Consumer Assistance at its toll-free number (800) 638-2041 or (301) 796-7100 or at its Internet address <http://www.fda.gov/MedicalDevices/ResourcesforYou/Industry/default.htm>.

Sincerely yours,

for 

Malvina B. Eydelman, M.D.
Director
Division of Ophthalmic, Neurological, and Ear,
Nose and Throat Devices
Office of Device Evaluation
Center for Devices and Radiological Health

Enclosure

Attachment E

Indications For Use Statement

Device Name: Alden HP 54 Spherical Lenses
 Alden HP 54 Toric Lenses
 Alden HP 54 Multifocal Lenses
 Alden HP 54 Toric Multifocal Lenses

Indications for Use: The Alden HP 54 Spherical soft contact lens is indicated for the correction of visual acuity in aphakic and not aphakic persons with non-diseased eyes with refractive ametropia (myopia or hyperopia). The lens may be worn by persons who exhibit refractive astigmatism of 1.50 diopters or less where the astigmatism does not interfere with visual acuity.

The Alden HP 54 Toric soft contact lens is indicated for the correction of visual acuity in aphakic and not aphakic persons with non-diseased eyes, with refractive ametropia (myopia or hyperopia), and/or possess refractive astigmatism not exceeding 10.00 diopters.

The Alden HP 54 Multifocal soft contact lens is indicated for the correction of visual acuity in aphakic and not aphakic persons with non-diseased eyes with refractive ametropia (myopia or hyperopia) and presbyopia, with add powers not exceeding 4.00 diopters. The lens may be worn by persons who exhibit refractive astigmatism of 1.50 diopters or less where the astigmatism does not interfere with visual acuity.

The Alden HP 54 Toric Multifocal soft contact lens is indicated for the correction of visual acuity in aphakic and not aphakic persons with non-diseased eyes, with refractive ametropia (myopia or hyperopia), and/or possess refractive astigmatism not exceeding 10.00 diopters, and presbyopia with add powers not exceeding 4.00 diopters.

The lenses are available for either conventional wear or planned replacement modalities.

(PLEASE DO NOT WRITE BELOW THIS LINE - CONTINUE ON ANOTHER PAGE IF NEEDED)

Concurrence of CDRH, Office of Device Evaluation (ODE)

Prescription Use
 (Per 21 CFR 801.109)

or

Over-The-Counter Use

(Optional Format 1-2-96)

Karen Warkentin

(Division Sign-Off)
 Division of Ophthalmic Devices

510(k) Number

K091327