

1091385

510(k) Summary

Submitted By:

Karen Bradburn, RAC
Senior Regulatory Affairs Specialist
Cook Incorporated
750 Daniels Way, PO Box 489
Bloomington, IN 47402

AUG 07 2009

Device:

Trade Name:	Approach Hydro ST Wire Guide
Proposed Classification:	Wire, Guide, Catheter
	21 CFR §870.1330

Indications for Use:

The Approach Hydro ST Wire Guide is indicated for use in facilitating delivery of percutaneous catheters into the peripheral vasculature.

Predicate Devices:

The Approach Hydro ST Wire Guide is identical in terms of intended use and similar in terms of materials of construction and technological characteristics to predicate devices reviewed as devices for facilitating delivery of percutaneous catheters into the peripheral vasculature.

Device Description:

The Approach Hydro ST Wire Guide is manufactured using a stainless steel wire with a PTFE coating and a stainless steel and platinum distal tip. The wire guide has a hydrophilic coating. The maximum outside diameter is 0.0142 inches and will be available in 135, 190 and 300 centimeter lengths. It will be supplied sterile, intended for one-time use.

Substantial Equivalence:

The Approach Hydro ST Wire Guide is similar to many devices in commercial distribution for facilitating delivery of percutaneous catheters into the peripheral vasculature. The identical indications for use, principles of operations, similar materials of construction and technological characteristics of the wire guide support a determination of substantial equivalence.

Test Data:

The Approach Hydro ST Wire Guide was subjected to the following tests to assure reliable design and performance under the specified testing parameters. These tests were comprised of:

1. Tensile Test
2. Tip Stiffness
3. Fracture Test
4. Flexing Test
5. Torque Strength Test
6. Torque Response Test
7. Lubricity Test
8. Corrosion Resistance Test
9. Biocompatibility Testing
10. Bioburden Testing
11. Endotoxin Testing
12. EtO Residual Testing

The results of these tests provide reasonable assurance that the device has been designed and tested to assure conformance to the requirements for its use as a wire guide.



Food and Drug Administration
10903 New Hampshire Avenue
Document Control Room - WO66-0609
Silver Spring, MD 20993-0002

AUG 07 2009

Cook Incorporated
% Ms. Karen Bradburn, RAC
Senior Regulatory Affairs Specialist
750 Daniels Way
P.O. Box 489
Bloomington, Indiana 47402-0489

Re: **K091385**

Trade/Device Name: Approach Hydro ST Wire Guide
Regulation Number: 21 CFR 870.1330
Regulation Name: Catheter Guide Wire
Regulatory Class: II
Product Code: DQX
Dated: July 7, 2009
Received: July 8, 2009

Dear Ms. Bradburn:

We have reviewed your Section 510(k) premarket notification of intent to market the device referenced above and have determined the device is substantially equivalent (for the indications for use stated in the enclosure) to legally marketed predicate devices marketed in interstate commerce prior to May 28, 1976, the enactment date of the Medical Device Amendments, or to devices that have been reclassified in accordance with the provisions of the Federal Food, Drug, and Cosmetic Act (Act) that do not require approval of a premarket approval application (PMA). You may, therefore, market the device, subject to the general controls provisions of the Act. The general controls provisions of the Act include requirements for annual registration, listing of devices, good manufacturing practice, labeling, and prohibitions against misbranding and adulteration.

If your device is classified (see above) into either class II (Special Controls) or class III (PMA), it may be subject to additional controls. Existing major regulations affecting your device can be found in the Code of Federal Regulations, Title 21, Parts 800 to 898. In addition, FDA may publish further announcements concerning your device in the Federal Register.

Please be advised that FDA's issuance of a substantial equivalence determination does not mean that FDA has made a determination that your device complies with other requirements of the Act or any Federal statutes and regulations administered by other Federal agencies.

Page 2-Ms. Karen Bradburn

You must comply with all the Act's requirements, including, but not limited to: registration and listing (21 CFR Part 807); labeling (21 CFR Part 801); medical device reporting (reporting of medical device-related adverse events) (21 CFR 803); good manufacturing practice requirements as set forth in the quality systems (QS) regulation (21 CFR Part 820); and if applicable, the electronic product radiation control provisions (Sections 531-542 of the Act); 21 CFR 1000-1050.

If you desire specific advice for your device on our labeling regulation (21 CFR Part 801), please go to


<http://www.fda.gov/AboutFDA/CentersOffices/CDRH/CDRHOffices/ucm115809.htm> for the Center for Devices and Radiological Health's (CDRH's) Office of Compliance. Also, please note the regulation entitled, "Misbranding by reference to premarket notification" (21CFR Part 807.97). For questions regarding the reporting of adverse events under the MDR regulation (21 CFR Part 803), please go to

<http://www.fda.gov/MedicalDevices/Safety/ReportaProblem/default.htm> for the CDRH's Office of Surveillance and Biometrics/Division of Postmarket Surveillance.

You may obtain other general information on your responsibilities under the Act from the Division of Small Manufacturers, International and Consumer Assistance at its toll-free number (800) 638-2041 or (240) 276-3150 or at its Internet address

<http://www.fda.gov/MedicalDevices/ResourcesforYou/Industry/default.htm>.

Sincerely yours,



Bram Zuckerman, M.D.

Director

Division of Cardiovascular Devices

Office of Device Evaluation

Center for Devices and Radiological Health

Enclosure

510(k) Number (if known): K091385

Device Name: Approach Hydro ST Wire Guide

Indications for Use:

Indicated for use in facilitating delivery of percutaneous catheters into the peripheral vasculature.

Prescription Use X
(Per 21 CFR 801 Subpart D)

OR

Over-the-Counter Use _____
(21 CFR 807 Subpart C)

(PLEASE DO NOT WRITE BELOW THIS LINE-CONTINUE ON ANOTHER PAGE IF NEEDED)

Concurrence of CDRH, Office of Device Evaluation (ODE)

Diana R. Verheij

(Division Sign-Off)
Division of Cardiovascular Devices

510(k) Number K091385