

KO R 1390

HDC S.r.l.  
SPECIAL 510(k) Notification

Sterile Self Ligating Spider Screw

JUL 30 2009

**510(k) Summary for the**

**Sterile Self Ligating Spider Screw**

This Special 510(k) Summary is being submitted in accordance with the requirements of the SMDA 1990 and 21 CFR 807.92.

**General Information**

Submitter: HDC S.r.l.  
Via dell'Industria, 19  
36060 – SARCEDO (Vicenza) - Italy  
Tel: +39 0445 364 148  
Fax: +39 0445 383 645

Establishment Registration Number: 3005728057

Contact Person: Guido Bonapace (consultant)  
ISEMED srl  
Via Borgo Santa Cristina, 12  
40026 Imola (BO)  
Italy  
Mob.phone: +39-335-537 8686  
Telephone: +39-051-625 7315  
Fax: +39-051-628 4344  
Email: gbonapace@isemed.eu

Summary Preparation Date: April 30, 2009

**Names**

Device Name: Sterile Self Ligating Spider Screw  
Classification Name: Endosseous dental implant  
Product Code: OAT  
Regulation Number: 872.3640  
Device Class: Class II

**Device Description**

The Sterile Self Ligating Spider Screw is a titanium fixation device to be inserted into the upper or lower jaws, designed to be immediately used (after the bone insertion) as fixation for orthodontic appliances. It is used temporarily and must be removed after the orthodontic treatment has been completed. It is provided sterile and is intended for single use only. The Sterile Self Ligating Spider Screw is provided in "Self Drilling and Self Tapping" configuration and is manufactured from Grade 5 Titanium alloy according to ASTM F136-98 standard. Sterile Self Ligating Spider Screw is provided with different diameters, lengths and head shapes.

**Predicate Devices**

The Sterile Self Ligating Spider Screw is substantially equivalent to the following predicate device, which is legally marketed in the United States:

<i>Applicant</i>	<i>Device name</i>	<i>510(k) Number</i>
HDC S.r.l.	HDC Sterile Spider Screw	K071851

The components of Sterile Self Ligating Spider Screw have been compared in terms of intended use, design characteristics, and safety solutions with the predicate device. Complete substantial equivalence information is provided in the Section XI of this submission.

**Intended Use**

The Sterile Self Ligating Spider Screw is a threaded titanium dental implant screw, intended to serve as a fixed anchorage point for attachment of orthodontic appliances to facilitate the orthodontic movement of the teeth. It is used temporarily and must be removed after the orthodontic treatment has been completed. It is provided sterile and is intended for single use only.

The Indications for use are substantially unchanged compared with K071851. The only changes are relevant to the fixation instructions of the Sterile Self Ligating Spider Screw: these instructions are added to the predicate's one.

**Fundamental Scientific Technology**

The Sterile Self Ligating Spider Screw does not introduce changes compared with the predicate device with regards to the Fundamental Scientific Technology.



Food and Drug Administration  
9200 Corporate Boulevard  
Rockville MD 20850

HDC S.r.l.  
C/O Mr. Guido Bonapace  
Regulatory Consultant  
ISEMED S.r.l.  
Via Borgo Santa Cristina, 12  
40026 Imola (BO)  
ITALY

JUL 30 2009

Re: K091390  
Trade/Device Name: Sterile Self Ligating Spider Screw  
Regulation Number: 21 CFR 872.3640  
Regulation Name: Endosseous Dental Implant  
Regulatory Class: II  
Product Code: OAT  
Dated: June 30, 2009  
Received: July 2, 2009

Dear Mr. Bonapace:

We have reviewed your Section 510(k) premarket notification of intent to market the device referenced above and have determined the device is substantially equivalent (for the indications for use stated in the enclosure) to legally marketed predicate devices marketed in interstate commerce prior to May 28, 1976, the enactment date of the Medical Device Amendments, or to devices that have been reclassified in accordance with the provisions of the Federal Food, Drug, and Cosmetic Act (Act) that do not require approval of a premarket approval application (PMA). You may, therefore, market the device, subject to the general controls provisions of the Act. The general controls provisions of the Act include requirements for annual registration, listing of devices, good manufacturing practice, labeling, and prohibitions against misbranding and adulteration.

If your device is classified (see above) into either class II (Special Controls) or class III (PMA), it may be subject to additional controls. Existing major regulations affecting your device can be found in the Code of Federal Regulations, Title 21, Parts 800 to 898. In addition, FDA may publish further announcements concerning your device in the Federal Register.

Page 2- Mr. Bonapace

Please be advised that FDA's issuance of a substantial equivalence determination does not mean that FDA has made a determination that your device complies with other requirements of the Act or any Federal statutes and regulations administered by other Federal agencies. You must comply with all the Act's requirements, including, but not limited to: registration and listing (21 CFR Part 807); labeling (21 CFR Part 801); medical device reporting (reporting of medical device-related adverse events) (21 CFR 803); good manufacturing practice requirements as set forth in the quality systems (QS) regulation (21 CFR Part 820); and if applicable, the electronic product radiation control provisions (Sections 531-542 of the Act); 21 CFR 1000-1050.

If you desire specific advice for your device on our labeling regulation (21 CFR Part 801), please go to <http://www.fda.gov/AboutFDA/CentersOffices/CDRH/CDRHOffices/ucm115809.htm> for the Center for Devices and Radiological Health's (CDRH's) Office of Compliance. Also, please note the regulation entitled, "Misbranding by reference to premarket notification" (21CFR Part 807.97). For questions regarding the reporting of adverse events under the MDR regulation (21 CFR Part 803), please go to <http://www.fda.gov/cdrh/mdr/> for the CDRH's Office of Surveillance and Biometrics/Division of Postmarket Surveillance.

You may obtain other general information on your responsibilities under the Act from the Division of Small Manufacturers, International and Consumer Assistance at its toll-free number (800) 638-2041 or (240) 276-3150 or at its Internet address <http://www.fda.gov/cdrh/industry/support/index.html>.

Sincerely yours,



Susan Runner, D.D.S., M.A.  
Acting Division Director  
Division of Anesthesiology, General Hospital,  
Infection Control and Dental Devices  
Office of Device Evaluation  
Center for Devices and  
Radiological Health

Enclosure

**INDICATIONS FOR USE STATEMENT**

Special 510(k) Number (if known): K091390

Device Name: Sterile Self Ligating Spider Screw

Indications for Use:

The Sterile Self Ligating Spider Screw is a threaded titanium dental implant screw, intended to serve as a fixed anchorage point for attachment of orthodontic appliances to facilitate the orthodontic movement of the teeth. The device is used temporarily and shall be removed after orthodontic treatment has been completed. It is provided sterile and is intended for single use only.

Prescription Use X  
(Part 21 CFR 801 Subpart D)

AND/OR

Over-The-Counter Use \_\_\_\_\_  
(21 CFR 801 Subpart C)

(PLEASE DO NOT WRITE BELOW THIS LINE-CONTINUE ON ANOTHER PAGE IF NEEDED)

---

Concurrence of CDRH, Office of Device Evaluation (ODE)



(Division Sign-Off)

Division of Anesthesiology, General Hospital  
Infection Control, Dental Devices

510(k) Number: K091390