

K091392

510(k) Summary
Pou Yu Biotechnology Co., Ltd
TDS Titanium Abutment for Nobel Biocare Branemark

JUL 30 2009

ADMINISTRATIVE INFORMATION

Manufacturer Name: Pou Yu Biotechnology Co., Ltd.
No. 6 Fugong Rd.
Fusing Township
Changhua County 506, Taiwan
Telephone: +886-(0)4 768 5660 x5122
Fax: +886-(0)4 768 9032

Official Contact: Daniel Tsao

DEVICE NAME AND CLASSIFICATION

Trade/Proprietary Name: TDS Titanium Abutment for Nobel Biocare Branemark,
Common Name: Dental implant abutment
Classification Regulations: Endosseous dental implant abutment
Class II, 21 CFR 872.3630
Product Code: NHA
Classification Panel: Dental Products Panel
Reviewing Branch: Dental Devices Branch

INTENDED USE

TDS Titanium Abutment for Nobel Biocare Branemark is intended for use with dental implants as a support for single or multiple tooth prostheses in the maxilla or mandible of a partially or fully edentulous patient.

DEVICE DESCRIPTION

TDS Titanium Abutment for Nobel Biocare Branemark are titanium abutments made of Ti-6Al-4V ELI titanium designed to be used in conjunction with specific dental implants utilizing the TDS Abutment screw, which is also made of Ti-6Al-4V ELI titanium and is used to secure the abutment to the implant. In combination with the implant, the abutments support single or multi-unit cement-retained restorations in the maxillary and/or mandibular arch. TDS Titanium Abutment for Nobel Biocare Branemark is compatible with the following implant systems which have an external hex with flat-to-flat dimensions of 2.4mm or greater: Nobel Biocare Branemark, 3i, BioHorizons, and Lifecore.

EQUIVALENCE TO MARKETED DEVICE

Pou Yu Biotechnology Co. Ltd demonstrated that, for the purposes of FDA's regulation of medical devices, TDS Titanium Abutment for Nobel Biocare Branemark is substantially equivalent in indications and design principles to predicate devices, each of which has been determined by FDA to be substantially equivalent to preamendment devices.



Food and Drug Administration
9200 Corporate Boulevard
Rockville MD 20850

Mr. Daniel Tsao
Manager
Pou Yu Biotechnology Company, Limited
No. 6 Fugong Road
Fusing Township
Changhua County 506
TAIWAN

JUL 30 2009

Re: K091392

Trade/Device Name: TDS Titanium Abutment for Nobel Biocare Branemark

Regulation Number: 21 CFR 872.3630

Regulation Name: Preformed Cusp

Regulatory Class: II

Product Code: NHA

Dated: May 8, 2009

~~Received: May 21, 2009~~

Dear Mr. Tsao:

We have reviewed your Section 510(k) premarket notification of intent to market the device referenced above and have determined the device is substantially equivalent (for the indications for use stated in the enclosure) to legally marketed predicate devices marketed in interstate commerce prior to May 28, 1976, the enactment date of the Medical Device Amendments, or to devices that have been reclassified in accordance with the provisions of the Federal Food, Drug, and Cosmetic Act (Act) that do not require approval of a premarket approval application (PMA). You may, therefore, market the device, subject to the general controls provisions of the Act. The general controls provisions of the Act include requirements for annual registration, listing of devices, good manufacturing practice, labeling, and prohibitions against misbranding and adulteration.


If your device is classified (see above) into either class II (Special Controls) or class III (PMA), it may be subject to additional controls. Existing major regulations affecting your device can be found in the Code of Federal Regulations, Title 21, Parts 800 to 898. In addition, FDA may publish further announcements concerning your device in the Federal Register.

Please be advised that FDA's issuance of a substantial equivalence determination does not mean that FDA has made a determination that your device complies with other requirements of the Act or any Federal statutes and regulations administered by other Federal agencies. You must comply with all the Act's requirements, including, but not limited to: registration and listing (21 CFR Part 807); labeling (21 CFR Part 801); medical device reporting (reporting of medical device-related adverse events) (21 CFR 803); good manufacturing practice requirements as set forth in the quality systems (QS) regulation (21 CFR Part 820); and if applicable, the electronic product radiation control provisions (Sections 531-542 of the Act); 21 CFR 1000-1050.

If you desire specific advice for your device on our labeling regulation (21 CFR Part 801), please go to <http://www.fda.gov/AboutFDA/CentersOffices/CDRH/CDRHOffices/ucm115809.htm> for the Center for Devices and Radiological Health's (CDRH's) Office of Compliance. Also, please note the regulation entitled, "Misbranding by reference to premarket notification" (21CFR Part 807.97). For questions regarding the reporting of adverse events under the MDR regulation (21 CFR Part 803), please go to <http://www.fda.gov/cdrh/mdr/> for the CDRH's Office of Surveillance and Biometrics/Division of Postmarket Surveillance.

You may obtain other general information on your responsibilities under the Act from the Division of Small Manufacturers, International and Consumer Assistance at its toll-free number (800) 638-2041 or (240) 276-3150 or at its Internet address <http://www.fda.gov/cdrh/industry/support/index.html>.

Sincerely yours,



Susan Runner, D.D.S., M.A.

Acting Director

Division of Anesthesiology, General Hospital,

Infection Control and Dental Devices

Office of Device Evaluation

Center for Devices and

Radiological Health

Enclosure

501(k) Premarket Notification

TDS Titanium Abutment for Nobel Biocare Branemark

Indications for Use

510(k) Number (if known): K091392

Device Name: TDS Titanium Abutment for Nobel Biocare Branemark

Indications for Use:

TDS Titanium Abutment for Nobel Biocare Branemark is intended for use with dental implants as a support for single or multiple tooth prostheses in the maxilla or mandible of a partially or fully edentulous patient.

This device is compatible with the following implant systems which have an external hex with flat-to-flat dimensions of 2.4mm or greater: **Nobel Biocare:** Branemark System® Mk III Groovy, Branemark System® Mk III Shorty, Branemark System® Zygoma, NobelSpeedy Groovy, NobelSpeedy Shorty; **3i:** Nano Tite External Hex Connection Implants, Full OSSEOTITE External Hex Connection Implants, OSSEOTITE External Hex Connection Implant; **BioHorizons:** Maestro - External Hex Implants; **Lifecore:** Restore External Hex Implant Systems.

Prescription Use X
(Part 21 CFR 801 Subpart D)

AND/OR

Over-The-Counter Use _____
(21 CFR 801 Subpart C)

(PLEASE DO NOT WRITE BELOW THIS LINE-CONTINUE ON ANOTHER PAGE OF
NEEDED)

Concurrence of CDPH, Office of Device Evaluation (ODE)

Susan [Signature]
(Division Sign-Off)

Division of Anesthesiology, General Hospital
Infection Control, Dental Devices

Page 1 of 1

510(k) Number: K091392