

510(k) Summary

JUL - 8 2009

Trade Name: Azur™ Peripheral HydroCoil® Endovascular Embolization System – Pushable 18

Generic Name: Vascular Embolization Coil

Classification: Class II, 21 CFR 870.3300

Submitted By: MicroVention, Inc
75 Columbia
Aliso Viejo, California U.S.A.

Contact: Naomi Gong

Predicate Devices:

Number	Description	Clearance Date
K071939	Azur™ Peripheral HydroCoil® Endovascular Embolization System- Pushable 18	January 11, 2008
K050954	HydroCoil Embolic System	June 28, 2005

Device Description

The Azur™ Peripheral HydroCoil® Endovascular Embolization System- Pushable 18 coils consist of an implantable coil packaged in an introducer along with a stylet. The stainless steel stylet is used to deploy the coil from the introducer into the delivery catheter. The coil is delivered to the treatment site through the delivery catheter using a standard guidewire. The Azur™ coil is platinum-based with a hydrogel polymer layer.

Indication For Use

The intended use as stated in the product labeling is as follows:

The Azur Peripheral HydroCoil Endovascular Embolization Coil System is intended to reduce or block the rate of blood flow in vessels of the peripheral vasculature. It is intended for use in the interventional radiologic management of arteriovenous malformations, arteriovenous fistulae, aneurysms, and other lesions of the peripheral vasculature.

Verification and Test Summary Table

Bench Testing	Result
Visual Inspection	Met established criteria
Dimensional Measurement	Met established criteria
Simulated Use	Met established criteria
Advancement force	Met established criteria
Distal tip tensile test	Met established criteria
Proximal tip tensile test	Met established criteria
Body coil tensile test	Met established criteria

Summary of Substantial Equivalence

The data presented in this submission demonstrates the technological similarity and equivalency of the Azur™ Peripheral HydroCoil® Endovascular Embolization System – Pushable 18 coils when compared with the predicate devices, MicroVention Azur Pushable 18 coils [K071939] and HydroCoil Embolic System (HES) [K050954].

The devices,

- Have the same intended use,
- Use the same operating principle,
- Incorporate the same basic design,
- Use similar construction and material,
- Are packaged and sterilized using same material and processes.

In summary, the Azur Pushable 18 coils described in this submission is, in our opinion, substantially equivalent to the predicate device.



Food and Drug Administration
9200 Corporate Boulevard
Rockville MD 20850

JUL - 8 2009

MicroVention, Inc.
c/o Ms. Naomi Gong
Regulatory Affairs Project Manager
75 Columbia, Suite A
Alison Viejo, CA 92656

Re: K091882

Azur Peripheral HydroCoil Endovascular Embolization Coil System – Pushable 18
Regulation Number: 21 CFR 870.3300
Regulation Name: Vascular embolization device
Regulatory Class: Class II (two)
Product Code: KR D
Dated: June 22, 2009
Received: June 24, 2009

Dear Ms. Gong:

We have reviewed your Section 510(k) premarket notification of intent to market the device referenced above and have determined the device is substantially equivalent (for the indications for use stated in the enclosure) to legally marketed predicate devices marketed in interstate commerce prior to May 28, 1976, the enactment date of the Medical Device Amendments, or to devices that have been reclassified in accordance with the provisions of the Federal Food, Drug, and Cosmetic Act (Act) that do not require approval of a premarket approval application (PMA). You may, therefore, market the device, subject to the general controls provisions of the Act. The general controls provisions of the Act include requirements for annual registration, listing of devices, good manufacturing practice, labeling, and prohibitions against misbranding and adulteration.

If your device is classified (see above) into either class II (Special Controls) or class III (PMA), it may be subject to additional controls. Existing major regulations affecting your device can be found in the Code of Federal Regulations, Title 21, Parts 800 to 898. In addition, FDA may publish further announcements concerning your device in the Federal Register.


Page 2 - Ms. Naomi Gong

Please be advised that FDA's issuance of a substantial equivalence determination does not mean that FDA has made a determination that your device complies with other requirements of the Act or any Federal statutes and regulations administered by other Federal agencies. You must comply with all the Act's requirements, including, but not limited to: registration and listing (21 CFR Part 807); labeling (21 CFR Part 801); medical device reporting (reporting of medical device-related adverse events) (21 CFR 803); good manufacturing practice requirements as set forth in the quality systems (QS) regulation (21 CFR Part 820); and if applicable, the electronic product radiation control provisions (Sections 531-542 of the Act); 21 CFR 1000-1050.

If you desire specific advice for your device on our labeling regulation (21 CFR Part 801), please go to <http://www.fda.gov/cdrh/comp/> for the Center for Devices and Radiological Health's (CDRH's) Office of Compliance. Also, please note the regulation entitled, "Misbranding by reference to premarket notification" (21CFR Part 807.97). For questions regarding the reporting of adverse events under the MDR regulation (21 CFR Part 803), please go to <http://www.fda.gov/cdrh/mdr/> for the CDRH's Office of Surveillance and Biometrics/Division of Postmarket Surveillance.

You may obtain other general information on your responsibilities under the Act from the Division of Small Manufacturers, International and Consumer Assistance at its toll-free number (800) 638-2041 or (240) 276-3150 or at its Internet address <http://www.fda.gov/cdrh/industry/support/index.html>.

Sincerely yours,



 Bram D. Zuckerman, M.D.
Director
Division of Cardiovascular Devices
Office of Device Evaluation
Center for Devices and
Radiological Health

Enclosure

Indications for Use

510(k) Number (if known): K091882

Device Name: Azur™ Peripheral HydroCoil® Endovascular Embolization System – Pushable 18

Indications For Use:

The Azur Peripheral HydroCoil Endovascular Embolization Coil System is intended to reduce or block the rate of blood flow in vessels of the peripheral vasculature. It is intended for use in the interventional radiologic management of arteriovenous malformations, arteriovenous fistulae, aneurysms, and other lesions of the peripheral vasculature.

Prescription Use X
(Part 21 CFR 801 Subpart D)

AND/OR

Over-The-Counter Use _____
(21 CFR 807 Subpart C)

(PLEASE DO NOT WRITE BELOW THIS LINE-CONTINUE ON ANOTHER PAGE IF NEEDED)

Concurrence of CDRH, Office of Device Evaluation (ODE)

Sumner R. Vachon
(Division Sign-Off)
Division of Cardiovascular Devices
510(k) Number K091882