

510(k) Summary - S8 Aspen

Date Prepared 25th June, 2009

Official Contact Dr Lionel King
V.P., Quality Assurance & Regulatory Affairs
ResMed Ltd
1 Elizabeth Macarthur Drive
Bella Vista NSW 2153
Australia

Tel: +61 (2) 8884 2243
Fax: +61 (2) 8884 2000

Classification Reference 21 CFR 868.5905

Product Code 73 BZD

Common/Usual Name Non continuous ventilator (IPPB).

Proprietary Name S8 Aspen

Predicate Device(s) VPAP ST (K080131)
VPAP Auto (K082605)

Reason for submission New Device

Indication for Use

The S8 Aspen self-adjusting system is indicated for the treatment of obstructive sleep apnea (OSA) in patients weighing more than 66 lb (>30 kg).

The S8 Aspen self-adjusting system is intended for home and hospital use.

Substantial Equivalence

The new device has the following similarities to the previously cleared predicate devices.

- Same intended use
- Similar operating principle
- Similar technologies
- Same manufacturing process

Design and Verification activities were performed on the S8 Aspen System as a result of the risk analysis and design requirements. All tests confirmed the product met the predetermined acceptance criteria. ResMed has determined that the new device has not altered the safety and effectiveness of CPAP treatment for patients with Obstructive Sleep Apnoea (OSA) who weigh more than 66 lb (>30 kg). The new device complies with the applicable requirements referenced in the FDA guidance documents:

- FDA Draft Reviewer Guidance for Ventilators (July 1995)
- FDA Guidance for the Content of Premarket Submissions for Software Contained in Medical Devices (May 11, 2005)

Device Description

The S8 Aspen System (S8 Aspen *with* HumidAire 4i Plus) is similar to the predicate device(s), with a new and improved power supply and heater controller for the HumidAire 4i Plus. The S8 Aspen contains a Micro-processor controlled blower system that generates Continuous Positive Airway Pressure (CPAP) from 4-20 cmH₂O as required to maintain an "air splint" for effective treatment of OSA. The system comprises the flow generator, patient tubing, mask (patient interface) and humidifier.

The functional characteristics of the S8 Aspen system includes all the clinician and user friendly features of the predicate devices.

Conclusion

The S8 Aspen is substantially equivalent to the Predicate devices.



SEP 28 2009

Food and Drug Administration
10903 New Hampshire Avenue
Document Control Room - WO66-0609
Silver Spring, MD 20993-0002

ResMed Limited
C/O Mr. David D'Cruz
Vice President, Clinical & Regulatory Affairs
ResMed Corporation
9001 Spectrum Center Boulevard
Kearny Mesa, California 92123

Re: K091947

Trade/Device Name: S8 Aspen
Regulation Number: 21 CFR 868.5905
Regulation Name: Noncontinuous Ventilator (IPPB)
Regulatory Class: II
Product Code: BZD
Dated: June 25, 2009
Received: June 30, 2009

Dear Mr. D'Cruz:

We have reviewed your Section 510(k) premarket notification of intent to market the device referenced above and have determined the device is substantially equivalent (for the indications for use stated in the enclosure) to legally marketed predicate devices marketed in interstate commerce prior to May 28, 1976, the enactment date of the Medical Device Amendments, or to devices that have been reclassified in accordance with the provisions of the Federal Food, Drug, and Cosmetic Act (Act) that do not require approval of a premarket approval application (PMA). You may, therefore, market the device, subject to the general controls provisions of the Act. The general controls provisions of the Act include requirements for annual registration, listing of devices, good manufacturing practice, labeling, and prohibitions against misbranding and adulteration.

If your device is classified (see above) into either class II (Special Controls) or class III (PMA), it may be subject to additional controls. Existing major regulations affecting your device can be found in the Code of Federal Regulations, Title 21, Parts 800 to 898. In addition, FDA may publish further announcements concerning your device in the Federal Register.


Page 2- Mr. D'Cruz

Please be advised that FDA's issuance of a substantial equivalence determination does not mean that FDA has made a determination that your device complies with other requirements of the Act or any Federal statutes and regulations administered by other Federal agencies. You must comply with all the Act's requirements, including, but not limited to: registration and listing (21 CFR Part 807); labeling (21 CFR Part 801); medical device reporting (reporting of medical device-related adverse events) (21 CFR 803); good manufacturing practice requirements as set forth in the quality systems (QS) regulation (21 CFR Part 820); and if applicable, the electronic product radiation control provisions (Sections 531-542 of the Act); 21 CFR 1000-1050.

If you desire specific advice for your device on our labeling regulation (21 CFR Part 801), please go to <http://www.fda.gov/AboutFDA/CentersOffices/CDRH/CDRHOffices/ucm115809.htm> for the Center for Devices and Radiological Health's (CDRH's) Office of Compliance. Also, please note the regulation entitled, "Misbranding by reference to premarket notification" (21CFR Part 807.97). For questions regarding the reporting of adverse events under the MDR regulation (21 CFR Part 803), please go to <http://www.fda.gov/MedicalDevices/Safety/ReportaProblem/default.htm> for the CDRH's Office of Surveillance and Biometrics/Division of Postmarket Surveillance.

You may obtain other general information on your responsibilities under the Act from the Division of Small Manufacturers, International and Consumer Assistance at its toll-free number (800) 638-2041 or (301) 796-7100 or at its Internet address <http://www.fda.gov/MedicalDevices/ResourcesforYou/Industry/default.htm>.

Sincerely yours,

A handwritten signature in black ink, appearing to read "Susan Runner", followed by the word "for" in a smaller, cursive script.

Susan Runner, D.D.S., M.A.
Acting Division Director
Division of Anesthesiology, General Hospital,
Infection Control and Dental Devices
Office of Device Evaluation
Center for Devices and
Radiological Health

Enclosure

Indication for Use

510(k) Number (if known):

Device Name: S8 Aspen

Indication for Use

The S8 Aspen self-adjusting system is indicated for the treatment of obstructive sleep apnea (OSA) in patients weighing more than 66 lb (>30 kg).

The S8 Aspen self-adjusting system is intended for home and hospital use.

Prescription Use AND/OR

(Part 21 CFR 801 Subpart D)

Over-The-Counter Use

(Part 21 CFR 807 Subpart C)

(PLEASE DO NOT WRITE BELOW THIS LINE - CONTINUE ON ANOTHER PAGE IF NEEDED)

Concurrence of CDRH; Office of Device Evaluation (ODE)

[Signature] for L. Schultze
(Division Sign-Off)

Page 1 of 1

Division of Anesthesiology, General Hospital
Infection Control, Dental Devices

510(k) Number: K091947