

K092303

Merit Medical Systems, Inc.
Merit Hydrophilic Guide Wire
Traditional Premarket Notification 510(k)

Section 5
510(k) Summary

Section 5

510(k) Summary

OCT 27 2009

General Provisions

Submitter Name: Merit Medical Systems, Inc.
Address: 1600 West Merit Parkway
South Jordan, UT 84095
Telephone Number: (801) 208-4789
Fax Number: (801) 253-6919
Contact Person: Susan Christensen
Date of Preparation: July 28, 2009
Registration Number: 1721504

Subject Device

Trade Name: Merit Hydrophilic Guide Wire
Common/Usual Name: Guide Wire
Classification Name: Catheter Guide Wire

Predicate Devices

Trade Name: Radiofocus® Glidewire®
Classification Name: Catheter Guide Wire
Premarket Notification: K863138
Manufacturer: Terumo Medical Corporation
Trade Name: Hydrophilic Coated Guide Wire
Classification Name: Catheter Guide Wire
Premarket Notification: K000011
Manufacturer: Lake Region Medical

Classification

Class II
21 CFR § 870.1330, 74 DQX
Division of Cardiovascular Devices

Intended Use

The Merit Hydrophilic Guide Wire is intended to facilitate the placement of devices during diagnostic and interventional procedures.

Device Description	The Merit Hydrophilic Guide Wire consists of a jacketed core wire with a hydrophilic coating applied to the jacket. The wire will be offered with straight and angled tip configurations in various lengths.
Technological Characteristics	Technological characteristics of the subject Merit Hydrophilic Guide Wire are substantially equivalent to those of the predicate, the currently marketed Terumo Radiofocus® Glidewire® [K863138], and the Lake Region Medical Hydrophilic Coated Guide Wire [K000011].
Safety & Performance Tests	No performance standards have been established under Section 514 of the Food, Drug and Cosmetic Act for these devices. However, a battery of tests was performed according to protocols based on the requirements of industry standards and guidances and were shown to meet the acceptance criteria that were determined to demonstrate the safety and efficacy of the device.
Summary of Substantial Equivalence	Based on the indications for use, design, safety, and performance testing, the subject Merit Hydrophilic Guide Wire meets the requirements that are considered essential for its intended use and is substantially equivalent to the predicate devices, the currently marketed Radiofocus® Glidewire® manufactured by Terumo Medical Corporation and the Hydrophilic Coated Guide Wire manufactured by Lake Region Medical.



Food and Drug Administration
10903 New Hampshire Avenue
Document Control Room W-O66-0609
Silver Spring, MD 20993-0002

OCT 27 2009

Merit Medical Systems, Inc.
c/o Ms. Susan Christensen
1600 West Merit Parkway
South Jordan, UT 84095

Re: K092303
Trade/Device Name: Merit Hydrophilic Guide Wire
Common Name: Catheter Guidewire
Regulation Number: 21 CFR 870.1330
Regulatory Class: II
Product Code: DQX
Dated: July 28, 2009
Received: July 29, 2009

Dear Ms. Christensen:

We have reviewed your Section 510(k) premarket notification of intent to market the device referenced above and have determined the device is substantially equivalent (for the indications for use stated in the enclosure) to legally marketed predicate devices marketed in interstate commerce prior to May 28, 1976, the enactment date of the Medical Device Amendments, or to devices that have been reclassified in accordance with the provisions of the Federal Food, Drug, and Cosmetic Act (Act) that do not require approval of a premarket approval application (PMA). You may, therefore, market the device, subject to the general controls provisions of the Act. The general controls provisions of the Act include requirements for annual registration, listing of devices, good manufacturing practice, labeling, and prohibitions against misbranding and adulteration.

If your device is classified (see above) into either class II (Special Controls) or class III (PMA), it may be subject to additional controls. Existing major regulations affecting your device can be found in the Code of Federal Regulations, Title 21, Parts 800 to 898. In addition, FDA may publish further announcements concerning your device in the Federal Register.

Please be advised that FDA's issuance of a substantial equivalence determination does not mean that FDA has made a determination that your device complies with other requirements of the Act or any Federal statutes and regulations administered by other Federal agencies. You must comply with all the Act's requirements, including, but not limited to: registration and listing (21 CFR Part 807); labeling (21 CFR Part 801); medical device reporting (reporting of medical device-related adverse events) (21 CFR 803); good manufacturing practice requirements as set


Page 2 – Ms. Susan Christensen

forth in the quality systems (QS) regulation (21 CFR Part 820); and if applicable, the electronic product radiation control provisions (Sections 531-542 of the Act); 21 CFR 1000-1050.

If you desire specific advice for your device on our labeling regulation (21 CFR Part 801), please go to <http://www.fda.gov/AboutFDA/CentersOffices/CDRH/CDRHOffices/ucm115809.htm> for the Center for Devices and Radiological Health's (CDRH's) Office of Compliance. Also, please note the regulation entitled, "Misbranding by reference to premarket notification" (21CFR Part 807.97). For questions regarding the reporting of adverse events under the MDR regulation (21 CFR Part 803), please go to <http://www.fda.gov/MedicalDevices/Safety/ReportaProblem/default.htm> for the CDRH's Office of Surveillance and Biometrics/Division of Postmarket Surveillance.

You may obtain other general information on your responsibilities under the Act from the Division of Small Manufacturers, International and Consumer Assistance at its toll-free number (800) 638-2041 or (301) 796-7100 or at its Internet address <http://www.fda.gov/MedicalDevices/ResourcesforYou/Industry/default.htm>.

Sincerely yours,



~~For~~ Bram D. Zuckerman, M.D.
Director
Division of Cardiovascular Devices
Office of Device Evaluation
Center for Devices and
Radiological Health

Enclosure

Section 4

Indications for Use Statement

510(k) Number (if known): K092303

Device Name: Merit Hydrophilic Guide Wire

Indications for Use:

The Merit Hydrophilic Guide Wire is intended to facilitate the placement of devices during diagnostic and interventional procedures.

Prescription Use X
(Part 21 CFR 801 Subpart D)

AND/OR

Over-The-Counter Use _____
(21 CFR 801 Subpart C)

(PLEASE DO NOT WRITE BELOW THIS LINE—CONTINUE ON ANOTHER PAGE IF NEEDED)

Concurrence of CDRH, Office of Device Evaluation (ODE)



(Division Sign-Off)
Division of Cardiovascular Devices

510(k) Number K092303