

1092368

II. SUMMARY AND CERTIFICATION

A. 510(k) Summary

NOV - 3 2009

Submitter: SterilMed, Inc.
Contact Person: Garrett Ahlborg
11400 73rd Avenue North
Maple Grove, MN 55369
Ph: 763-488-3483
Fax: 763-488-2051
Date Prepared: August 3, 2009
Trade Name: Reprocessed Pulse Oximeter Sensors
Classification Name: Oximeter, Reprocessed
Classification Number: Class II, 21 CFR 870.2700
Product Code: NLF

Predicate Devices:	The SterilMed reprocessed pulse oximeter sensors are substantially equivalent to the Masimo LNCS [®] pulse oximeter sensors (K041815, K051212, & K060143) and the Masimo LNOP [®] pulse oximeter sensors (K031330).
Device Description:	<p>SterilMed's reprocessed Masimo LNCS[®] Pulse Oximeter Sensors consist of a sensor, integrated sensor cable, and the sensor plug which connects to the Pulse Oximeter. The reprocessed Masimo LNOP[®] Pulse Oximeter Sensors consist of a sensor (two LED's and a photodiode), connector extension and connector, but do not feature an integrated cable. Both configurations have a sensor that uses an optical means to determine the light absorption of functional arterial hemoglobin. The sensor contains three optical components: two light emitting diodes (LED's) that serve as light sources, and one photodiode, that acts as a light receiver. The oximeter sensor is positioned so that the LED's and photodiode oppose one another across the tissue. The sensor is connected via cable to a pulse oximeter, which provides continuous noninvasive, self-calibrated measurements of both oxygen saturation of functional hemoglobin and pulse rate.</p> <p>Note: Only the pulse oximeter sensor is the subject of this submission, the oximeter and any other related equipment are not included in the scope of this submission.</p>
Intended Use:	The reprocessed pulse oximeter sensors are indicated for use for continuous noninvasive arterial oxygen saturation (SpO ₂) and pulse rate monitoring.
Functional and Safety Testing:	Representative samples of reprocessed pulse oximeter sensors were tested to demonstrate appropriate functional characteristics by utilizing the necessary bench testing and an <i>in vivo</i> clinical validation. Process validation testing was performed to validate the sterilization procedures as well as device packaging. In addition, the manufacturing process includes visual and validated functional testing of all products produced.
Conclusion:	<p>The reprocessed pulse oximeter sensors are substantially equivalent to the Masimo pulse oximeter sensors.</p> <p>This conclusion is based upon the devices' similarities in functional design (principles of operation), materials, indications for use and methods of construction.</p>



Food and Drug Administration
10903 New Hampshire Avenue
Document Control Room W-066-0609
Silver Spring, MD 20993-0002

Mr. Garrett Ahlborg
Regulatory Affairs Specialist
SterilMed, Incorporated
11400 73rd Avenue North, Suite 100
Minneapolis, Minnesota 55369

NOV - 3 2009

Re: K092368
Trade/Device Name: Reprocessed Pulse Oximeter Sensors
Regulation Number: 21 CFR 870.2700
Regulation Name: Oximeter
Regulatory Class: II
Product Code: NLF
Dated: August 3, 2009
Received: August 5, 2009

Dear Mr. Ahlborg:

We have reviewed your Section 510(k) premarket notification of intent to market the device referenced above and have determined the device is substantially equivalent (for the indications for use stated in the enclosure) to legally marketed predicate devices marketed in interstate commerce prior to May 28, 1976, the enactment date of the Medical Device Amendments, or to devices that have been reclassified in accordance with the provisions of the Federal Food, Drug, and Cosmetic Act (Act) that do not require approval of a premarket approval application (PMA). You may, therefore, market the device, subject to the general controls provisions of the Act. The general controls provisions of the Act include requirements for annual registration, listing of devices, good manufacturing practice, labeling, and prohibitions against misbranding and adulteration.


If your device is classified (see above) into either class II (Special Controls) or class III (PMA), it may be subject to additional controls. Existing major regulations affecting your device can be found in the Code of Federal Regulations, Title 21, Parts 800 to 898. In addition, FDA may publish further announcements concerning your device in the Federal Register.

Please be advised that FDA's issuance of a substantial equivalence determination does not mean that FDA has made a determination that your device complies with other requirements of the Act or any Federal statutes and regulations administered by other Federal agencies. You must comply with all the Act's requirements, including, but not limited to: registration and listing (21 CFR Part 807); labeling (21 CFR Part 801); medical device reporting (reporting of medical device-related adverse events) (21 CFR 803); good manufacturing practice requirements as set forth in the quality systems (QS) regulation (21 CFR Part 820); and if applicable, the electronic product radiation control provisions (Sections 531-542 of the Act); 21 CFR 1000-1050.

If you desire specific advice for your device on our labeling regulation (21 CFR Part 801), please go to <http://www.fda.gov/AboutFDA/CentersOffices/CDRH/CDRHOffices/ucm115809.htm> for the Center for Devices and Radiological Health's (CDRH's) Office of Compliance. Also, please note the regulation entitled, "Misbranding by reference to premarket notification" (21CFR Part 807.97). For questions regarding the reporting of adverse events under the MDR regulation (21 CFR Part 803), please go to <http://www.fda.gov/cdrh/mdr/> for the CDRH's Office of Surveillance and Biometrics/Division of Postmarket Surveillance.

You may obtain other general information on your responsibilities under the Act from the Division of Small Manufacturers, International and Consumer Assistance at its toll-free number (800) 638-2041 or (240) 276-3150 or at its Internet address <http://www.fda.gov/cdrh/industry/support/index.html>.

Sincerely yours,



Susan Runner, D.D.S., M.A.
Acting Division Director
Division of Anesthesiology, General Hospital,
Infection Control and Dental Devices
Office of Device Evaluation
Center for Devices and
Radiological Health

Enclosure

Indications for Use

510(k) Number (if known):

Device Name: Reprocessed Pulse Oximeter Sensors

Indications for Use:

The reprocessed pulse oximeter sensors are indicated for use for continuous noninvasive arterial oxygen saturation (SpO₂) and pulse rate monitoring.

Prescription Use X
(Part 21 CFR 801 Subpart D)

AND/OR

Over-The-Counter Use _____
(21 CFR 807 Subpart C)

(PLEASE DO NOT WRITE BELOW THIS LINE-CONTINUE ON ANOTHER PAGE IF NEEDED)

Concurrence of CDRH, Office of Device Evaluation (ODE)



(Division Sign-Off)

Division of Anesthesiology, General Hospital
Infection Control, Dental Devices

510(k) Number: K092368

Page 1 of 1