



Food and Drug Administration
10903 New Hampshire Avenue
Document Mail Center-WO66-G609
Silver Spring, MD 20993-0002

The Binding Site Ltd.
c/o Mr. Jay H. Geller
Authorized U.S. Representative
12100 Wilshire Boulevard, Suite 500
Los Angeles, CA 90025-7121

DEC 28 2009

Re: k092450

Trade/Device Name: Human IgD Kit for use on SPA_{PLUS}TM
Regulation Number: 21 CFR §866.5510
Regulation Name: Immunoglobulins A, G, M, D, E Immunological test system
Regulatory Class: II
Product Code: CZJ
Dated: November 12, 2009
Received: December 1, 2009

Dear Mr. Geller:

We have reviewed your Section 510(k) premarket notification of intent to market the device referenced above and have determined the device is substantially equivalent (for the indications for use stated in the enclosure) to legally marketed predicate devices marketed in interstate commerce prior to May 28, 1976, the enactment date of the Medical Device Amendments, or to devices that have been reclassified in accordance with the provisions of the Federal Food, Drug, and Cosmetic Act (Act) that do not require approval of a premarket approval application (PMA). You may, therefore, market the device, subject to the general controls provisions of the Act. The general controls provisions of the Act include requirements for annual registration, listing of devices, good manufacturing practice, labeling, and prohibitions against misbranding and adulteration.

If your device is classified (see above) into class II (Special Controls), it may be subject to such additional controls. Existing major regulations affecting your device can be found in Title 21, Code of Federal Regulations (CFR), Parts 800 to 895. In addition, FDA may publish further announcements concerning your device in the Federal Register.


Please be advised that FDA's issuance of a substantial equivalence determination does not mean that FDA has made a determination that your device complies with other requirements of the Act or any Federal statutes and regulations administered by other Federal agencies. You must comply with all the Act's requirements, including, but not limited to: registration and listing (21 CFR Part 807); labeling (21 CFR Parts 801 and 809); medical device reporting (reporting of medical device-related adverse events) (21 CFR 803); and good manufacturing practice requirements as set forth in the quality systems (QS) regulation (21 CFR Part 820). This letter

will allow you to begin marketing your device as described in your Section 510(k) premarket notification. The FDA finding of substantial equivalence of your device to a legally marketed predicate device results in a classification for your device and thus, permits your device to proceed to the market.

If you desire specific advice for your device on our labeling regulation (21 CFR Parts 801 and 809), please contact the Office of *In Vitro* Diagnostic Device Evaluation and Safety at (301) 796-5450. Also, please note the regulation entitled, "Misbranding by reference to premarket notification" (21 CFR Part 807.97). For questions regarding the reporting of adverse events under the MDR regulation (21 CFR Part 803), please go to <http://www.fda.gov/MedicalDevices/Safety/ReportaProblem/default.htm> for the CDRH's Office of Surveillance and Biometrics/Division of Postmarket Surveillance.

You may obtain other general information on your responsibilities under the Act from the Division of Small Manufacturers, International and Consumer Assistance at its toll-free number (800) 638-2041 or (301) 796-7100 or at its Internet address <http://www.fda.gov/cdrh/industry/support/index.html>.

Sincerely yours,



Maria M. Chan, Ph.D.

Director

Division of Immunology and Hematology Devices

Office of *In Vitro* Diagnostic Device Evaluation and Safety

Center for Devices and Radiological Health

Enclosure

Indications for Use

510(k) Number (if known): k 092450

Device Name: Human IgD Kit for use on SPAplus

Indications for Use: This kit is intended for measuring human immunoglobulin D (IgD) in serum as an aid in the diagnosis of abnormal protein metabolism and the body's lack of ability to resist infectious agents. The test result is to be used in conjunction with other clinical and laboratory findings.

Prescription Use X
(Part 21 CFR 801 Subpart D)

AND/OR

Over-The-Counter Use _____
(21 CFR 801 Subpart C)

(PLEASE DO NOT WRITE BELOW THIS LINE-CONTINUE ON ANOTHER PAGE
OF NEEDED)

Concurrence of CDRH, Office of Device Evaluation (ODE)

Leena Philip
Division Sign-Off

Attachment Q

Page 1 of 1

Office of In Vitro Diagnostic
Device Evaluation and Safety

58

510(k) k 092450