

SEP 15 2009

SECTION 8 - 510(K) SUMMARY

**510(k) SUMMARY OF SAFETY AND EFFECTIVENESS**  
**Force Fiber® Blue Polyethylene Non-Absorbable Surgical Sutures**

**A. Name, Address, Phone and Fax Number of Applicant**

Teleflex Medical Incorporated  
5307 95<sup>th</sup> Avenue  
Kenosha, WI 53144  
Phone: 262-925-8274  
Fax: 262-657-2801  
E-mail: jvoigt@teleflexmedical.com

**B. Contact Person**

Joy Voigt  
Regulatory Affairs Manager

**C. Date Prepared**

17 July 2009

**D. Device Name**

Trade Name:  
Force Fiber® Blue Ultra High Molecular Weight Polyethylene Non-Absorbable  
Surgical Suture

Common Name: Polyethylene Synthetic Non-Absorbable Surgical Suture

Classification Name: Nonabsorbable poly(ethylene terephthalate) surgical suture

**E. Device Description**

The Force Fiber® Polyethylene is non-absorbable, sterile, surgical suture composed of ultra high molecular weight polyethylene (UHMWPE). It is available as 100% blue (UHMWPE), sizes 0, 1 and 2 meeting USP requirements except for oversized diameter.

**F. Indications for Use**

Force Fiber® Blue Ultra High Molecular Weight Polyethylene Non-Absorbable Surgical Sutures are indicated for use in approximation and/or ligation of soft tissues, including use in cardiovascular surgeries and the use of allograft tissue for orthopaedic surgeries.

**G. Substantial Equivalence**

Teleflex Medical  
Force Fiber® Blue Ultra High Molecular Weight Polyethylene Non-Absorbable Surgical Suture  
Special PreMarket Notification (510(k)) Submission

The device has the same intended use and fundamental scientific technology as Teleflex Medical Force Fiber® Polyethylene Non-absorbable Surgical Suture (K063778). The determination of substantial equivalence for this device was based on a detailed device description, performance testing, and conformance with voluntary performance standards.

#### **H. Summary of Testing**

All sizes of Force Fiber® Blue Ultra High Molecular Weight Polyethylene Non-Absorbable Surgical Suture have been tested in accordance with USP 31 – Non-absorbable Surgical Sutures for Knot Pull Tensile Strength, Needle Attachment and Diameter, and meet the requirements of the *Class II Special Controls Guidance: Surgical Sutures*; Guidance for Industry and FDA; June 3, 2003.

All materials used in the fabrication of the Force Fiber® Blue Ultra High Molecular Weight Polyethylene Non-Absorbable Surgical Suture were evaluated through biological qualification safety tests as outlined in ISO 10993-1:2003; *Biological Evaluation of Medical Devices - Part 1: Evaluation and Testing*.



DEPARTMENT OF HEALTH & HUMAN SERVICES

Food and Drug Administration  
10903 New Hampshire Avenue  
Document Mail Center - WO66-G609  
Silver Spring, MD 20993-0002

SEP 15 2009

Teleflex Medical  
% Ms. Joy Voigt  
Manager, Regulatory Affairs  
5307 95<sup>th</sup> Avenue  
Kenosha, Wisconsin 53144

Re: K092533

Trade/Device Name: Force Fiber<sup>®</sup> Blue Ultra High Molecular Weight Polyethylene  
Non-Absorbable Surgical Suture

Regulation Number: 21 CFR 878.5000

Regulation Name: Nonabsorbable poly(ethylene terephthalate) surgical suture

Regulatory Class: Class II

Product Code: GAT

Dated: August 18, 2009

Received: August 19, 2009

Dear Ms. Voigt:

We have reviewed your Section 510(k) premarket notification of intent to market the device referenced above and have determined the device is substantially equivalent (for the indications for use stated in the enclosure) to legally marketed predicate devices marketed in interstate commerce prior to May 28, 1976, the enactment date of the Medical Device Amendments, or to devices that have been reclassified in accordance with the provisions of the Federal Food, Drug, and Cosmetic Act (Act) that do not require approval of a premarket approval application (PMA). You may, therefore, market the device, subject to the general controls provisions of the Act. The general controls provisions of the Act include requirements for annual registration, listing of devices, good manufacturing practice, labeling, and prohibitions against misbranding and adulteration.

If your device is classified (see above) into either class II (Special Controls) or class III (PMA), it may be subject to additional controls. Existing major regulations affecting your device can be found in the Code of Federal Regulations, Title 21, Parts 800 to 898. In addition, FDA may publish further announcements concerning your device in the Federal Register.

Please be advised that FDA's issuance of a substantial equivalence determination does not mean that FDA has made a determination that your device complies with other requirements of the Act or any Federal statutes and regulations administered by other Federal agencies. You must comply with all the Act's requirements, including, but not limited to: registration and listing (21 CFR Part 807); labeling (21 CFR Part 801); medical device reporting (reporting of medical device-related adverse events) (21 CFR 803); good manufacturing practice requirements as set


Page 2 - Ms. Joy Voigt

forth in the quality systems (QS) regulation (21 CFR Part 820); and if applicable, the electronic product radiation control provisions (Sections 531-542 of the Act); 21 CFR 1000-1050.

If you desire specific advice for your device on our labeling regulation (21 CFR Part 801), please go to <http://www.fda.gov/AboutFDA/CentersOffices/CDRH/CDRHOffices/ucm115809.htm> for the Center for Devices and Radiological Health's (CDRH's) Office of Compliance. Also, please note the regulation entitled, "Misbranding by reference to premarket notification" (21CFR Part 807.97). For questions regarding the reporting of adverse events under the MDR regulation (21 CFR Part 803), please go to <http://www.fda.gov/cdrh/mdr/> for the CDRH's Office of Surveillance and Biometrics/Division of Postmarket Surveillance.

You may obtain other general information on your responsibilities under the Act from the Division of Small Manufacturers, International and Consumer Assistance at its toll-free number (800) 638-2041 or (240) 276-3150 or at its Internet address <http://www.fda.gov/cdrh/industry/support/index.html>.

Sincerely yours,



Mark N. Melkersen  
Director  
Division of Surgical, Orthopedic  
and Restorative Devices  
Office of Device Evaluation  
Center for Devices and  
Radiological Health

Enclosure

1092533 11

Teleflex Medical  
Force Fiber® Blue Ultra High Molecular Weight Polyethylene Non-Absorbable Surgical Suture  
Special PreMarket Notification (510(k)) Submission

**SECTION 4 - INDICATIONS FOR USE STATEMENT**

510(k) Number (if known): \_\_\_\_\_

Device Name: Force Fiber® Blue Ultra High Molecular Weight Polyethylene  
Non-Absorbable Surgical Suture

**Indications for Use:**

Force Fiber® Polyethylene Non-Absorbable Surgical Sutures are indicated for use in approximation and/or ligation of soft tissues, including use in cardiovascular surgeries and the use of allograft tissue for orthopaedic surgeries.

Prescription Use  X   
(Part 21 CFR 801 Subpart D)

AND/OR

Over-The-Counter Use \_\_\_\_\_  
(21 CFR 801 Subpart C)

(PLEASE DO NOT WRITE BELOW THIS LINE-CONTINUE ON ANOTHER PAGE OF  
NEEDED)

Concurrence of CDRH, Office of Device Evaluation (ODE)

Daniel Krone for MKM   
(Division Sign-Off)  
Division of Surgical, Orthopedic,  
and Restorative Devices

Page \_\_\_ of \_\_\_

510(k) Number  1092533