

**Cachet Pharmaceuticals Pvt. Ltd.
Sterile Saline Solution
Premarket Submission 510 (K)**

**510(K) SUMMARY
FOR
Sterile Saline Solution**

SEP 10 2010

This summary is provided in accordance with the Safe Medical Devices Act (SMDA) of 1990. The information provided in the 510 (K), Premarket Notification, was in accordance with 21 CFR 807.87

1. Submitter of 510 (K)

Cachet Pharmaceuticals Pvt. Ltd.
415 Shah Nahar, Worli
Mumbai 400 018
India

Contact Person : Mr. Shivkumar Agrawal
Telephone No. : 91-22-4082-9999 or 2496-5831
Fax: 91-22-2497-3691
E.mail: skagrawal@cachetindia.com

Contact Person (United States): Mr. Haribabu Talasila
Telephone No. : 732-447-4353
E mail : htalasila@chemtexinternational.com.

**2. Product Code:
LPN**

3. Device Name
Classification Name: Soft (hydrophilic) Contact Lens Solution
Proprietary Name: Sterile Saline Solution

4. Legally Marketed Predicate Devices
Purilens Saline Solution – K002319
Lens Plus Rewetting Drops – K042562
Unisol -P7900 11

Cachet Pharmaceuticals Pvt. Ltd.
Sterile Saline Solution
Premarket Submission 510 (K)

5. Description of the Device

Sterile Saline Solution is a sterile, preservative free, buffered, isotonic, clear, colourless aqueous solution containing Sodium Chloride, Sodium Borate and Boric Acid.

The solution is packaged in plastic sterile single use twist open pods, reusable bottles with dropper tips and caps. Each bottle is closed with tamper resistant cap with a breakable ring, which must be broken in order to open the bottle.

6. Indications for Use

Sterile Saline Solution is indicated for rinsing & wetting of soft (hydrophilic) contact lenses.

7. Description of Safety and Substantial Equivalence:

A series of studies were completed to demonstrate the substantial equivalence of Sterile Saline Solution to the predicate device(s). All testing was conducted in accordance with and in conformance to applicable device regulations. Results demonstrate the solution is non-toxic and is comparable to other currently marketed soft contact lens solutions. Results from all tests demonstrate the substantial equivalence to previously approved predicate device(s).

Sterile Saline Solution is substantially equivalent in terms of its actions and indications for use, to Purilens Saline Solution –K002319, Lens Plus Rewetting Drops -K042562, Unisol- P790011 cleared for marketing under 510(K). Sterile Saline Solution meets the guideline set forth in FDA's May 1, 1997 Guidance for Industry, Premarket Notification 510 (K) Guidance Document for Contact Lens Care products.

Toxicity:

A series of cytotoxicity and eye irritation studies of the Sterile Solution were undertaken. In these studies, there was no evidence of toxicity.

Bacteriostasis Test:

A bacteriostasis study was conducted in accordance with Micro Appendix C of the Premarket Notification 510 (K) Guidance Document for Contact Lens Care Products. The purpose of the study is to evaluate the ability of bacteria to survive in the reclosable containers containing Sterile Saline Solution (unpreserved borate buffered saline). The results show that Staphylococcus aureus, Escherichia coli, Pseudomonas aeruginosa, Candida albicans and Aspergillus niger have no significant growth and survive over the designed period. The data support the desire discard statement on label up to 30 days.

Cachet Pharmaceuticals Pvt. Ltd.
Sterile Saline Solution
Premarket Submission 510 (K)

8. Substantial Equivalence

The data provide in this 510 (K) submission concludes that Sterile Saline Solution is substantially equivalent to currently cleared Purilens Saline Solution – K002319, Lens Plus Rewetting Drops –K042562, Unisol- P790011 cleared for marketing under 510 (K) for rinsing and wetting of soft (hydrophilic) contact lenses.

SUBSTANTIAL EQUIVALENCE CHART

Substantial Equivalency	Sterile Saline Solution	Purilens Saline Solution K002319	Lens Plus Rewetting Drops K042562	Unisol P790011
Manufacturer	Marck Bioscience Ltd.	Purilens Inc.	Advanced Medical Optics	Alcon Research Ltd.
Intended use (Indications for use)	For rinsing and wetting of soft (hydrophilic) contact lenses	For cleaning, disinfecting and storing of contact lenses	For moistening and dehydrating of contact lenses	For heat disinfecting rinsing, storage and wetting of soft (hydrophilic) contact lenses
Formulation	Sodium chloride, Sodium borate, Boric acid	Sodium chloride, Boric acid	Sodium chloride, Boric acid	Sodium chloride, Sodium borate, Boric acid
Preservative	No	No	No	No
Sterility claim	Sterile	Sterile	Sterile	Sterile



Food and Drug Administration
10903 New Hampshire Avenue
Document Control Room –WO66-G609
Silver Spring, MD 20993-0002

Cachet Pharmaceuticals Pvt., Ltd.
c/o Mr. Shivkumar Agrawal
415 Shah Nahar, Worli
Mumbai, India 400 018

SEP 10 2010

Re: K093367

Trade Name: Sterile Saline Solution
Regulation Number: 21 CFR 886.5928
Regulation Name: Soft (hydrophilic) contact lens care products
Regulatory Class: Class II
Product Code: LPN
Dated: September 10, 2010
Received: September 10, 2010

Dear Mr. Agrawal:

We have reviewed your Section 510(k) premarket notification of intent to market the device referenced above and have determined the device is substantially equivalent (for the indications for use stated in the enclosure) to legally marketed predicate devices marketed in interstate commerce prior to May 28, 1976, the enactment date of the Medical Device Amendments, or to devices that have been reclassified in accordance with the provisions of the Federal Food, Drug, and Cosmetic Act (Act) that do not require approval of a premarket approval application (PMA). You may, therefore, market the device, subject to the general controls provisions of the Act. The general controls provisions of the Act include requirements for annual registration, listing of devices, good manufacturing practice, labeling, and prohibitions against misbranding and adulteration. Please note: CDRH does not evaluate information related to contract liability warranties. We remind you, however, that device labeling must be truthful and not misleading.

If your device is classified (see above) into either class II (Special Controls) or class III (PMA), it may be subject to additional controls. Existing major regulations affecting your device can be found in the Code of Federal Regulations, Title 21, Parts 800 to 898. In addition, FDA may publish further announcements concerning your device in the Federal Register.

Please be advised that FDA's issuance of a substantial equivalence determination does not mean that FDA has made a determination that your device complies with other requirements of the Act or any Federal statutes and regulations administered by other Federal agencies. You must comply with all the Act's requirements, including, but not limited to: registration and listing (21 CFR Part 807); labeling (21 CFR Part 801); medical device reporting (reporting of medical device-related adverse events) (21 CFR 803); good manufacturing practice requirements as set forth in the quality systems (QS) regulation (21 CFR Part 820); and if applicable, the electronic product radiation control provisions (Sections 531-542 of the Act); 21 CFR 1000-1050.

If you desire specific advice for your device on our labeling regulation (21 CFR Part 801), please go to <http://www.fda.gov/AboutFDA/CentersOffices/CDRH/CDRHOffices/ucm115809.htm> for the Center for Devices and Radiological Health's (CDRH's) Office of Compliance. Also, please note the regulation entitled, "Misbranding by reference to premarket notification" (21CFR Part 807.97). For questions regarding the reporting of adverse events under the MDR regulation (21 CFR Part 803), please go to <http://www.fda.gov/MedicalDevices/Safety/ReportaProblem/default.htm> for the CDRH's Office of Surveillance and Biometrics/Division of Postmarket Surveillance.

You may obtain other general information on your responsibilities under the Act from the Division of Small Manufacturers, International and Consumer Assistance at its toll-free number (800) 638-2041 or (301) 796-7100 or at its Internet address <http://www.fda.gov/MedicalDevices/ResourcesforYou/Industry/default.htm>.

Sincerely yours,



Malvina B. Eydelman, M.D.
Director
Division of Ophthalmic, Neurological and Ear,
Nose and Throat Devices
Office of Device Evaluation
Center for Devices and
Radiological Health

Enclosure

K093367

Indications for Use

SEP 10 2010

510(k) Number (if known): K093367

Device Name: Sterile Saline Solution

Indications For Use:

Sterile Saline Solution is indicated for rinsing and wetting of soft (hydrophilic) contact lenses.

Prescription Use _____
(Part 21 CFR 801 Subpart D)

AND/OR

Over-The-Counter Use _____
(Part 21 CFR 801 Subpart C)

(PLEASE DO NOT WRITE BELOW THIS LINE-CONTINUE ON ANOTHER PAGE IF NEEDED)

Concurrence of CDRH, Office of Device Evaluation (ODE)

Denise Hauff, Ph.D.
(Division Sign-Off)
Division of Ophthalmic, Neurological and Ear,
Nose and Throat Devices

Page 1 of _____

510(k) Number K093367