

NOV 29 2010

**510(k) Summary for
INNOVANCE® D-Dimer Assay**

This summary of 510(k) safety and effectiveness information is being submitted in accordance with the requirements of SDMA 1990 and 21 CFR 807.92.

The assigned 510(k) number is: K093626

1. Manufacturer's Name, Address, Telephone, and Contact Person, Date of Preparation:

Manufacturer: Siemens Healthcare Diagnostics
Products GmbH
Emil-von-Behring Str. 76
D-35001, Marburg Germany

Contact Information: Siemens Healthcare Diagnostics Inc.
Glasgow Site
P.O. Box 6101
Newark, Delaware 19714
Attn: Kathleen Dray-Lyons
Tel: 781-826-4551
Fax: 781-826-2497

Preparation date: November 23, 2009

2. Device Name/ Classification: INNOVANCE® D-Dimer

Class: Fibrinogen and Fibrin Split Product, Antigen, Antiserum
and controls, Class II
21 CFR 864.7320
Panel: Hematology (HE)
Product Code: DAP

3. Identification of the Legally Marketed Device:

VIDAS® D-Dimer Exclusion™ – k040882

4. Device Description:

Polystyrene particles covalently coated with a monoclonal antibody (8D3) are aggregated when mixed with samples containing D-dimer. The D-dimer cross-linkage region has a stereosymmetrical structure, i.e. the epitope for the monoclonal antibody occurs twice. Consequently, one antibody suffices in order to trigger an aggregation reaction, which is then detected turbidimetrically via the increase in turbidity.

5. Device Intended Use:

INNOVANCE® D-Dimer:

For the quantitative determination of cross-linked fibrin degradation products (D-dimers) in human plasma on Siemens Healthcare Diagnostics and Sysmex® Coagulation Systems. The INNOVANCE® D-Dimer assay is intended for use in conjunction with a non-high clinical pretest probability (PTP) assessment model to exclude deep vein thrombosis (DVT) and pulmonary embolism (PE).

6. Medical device to which equivalence is claimed and comparison information:

The INNOVANCE® D-Dimer is substantially equivalent to the VIDAS® D-Dimer Exclusion™ (k040882) assay. The INNOVANCE® D-Dimer method, like the VIDAS® D-Dimer Exclusion™ method, is intended for use in conjunction with a clinical pretest probability (PTP) assessment model to exclude deep vein thrombosis (DVT) and pulmonary embolism (PE) disease.

7. Device Performance Characteristics:

Clinical Performance of the INNOVANCE® D-Dimer assay to exclude DVT

The INNOVANCE® D-Dimer assay was evaluated on the BCS® / BCS® XP System in a multi-center study to validate the exclusion of DVT using fresh specimens collected from 455 consecutive patients presenting to the emergency department with suspected DVT. Of these 455 patients, 29 were excluded for a total of 426 patients available for final analysis.

All patients were evaluated using the Wells' rules to estimate a likely or unlikely pre-test probability (PTP) of DVT. Patient specimens were tested with the INNOVANCE® D-Dimer assay and results were compared to a cutoff value of 0.50 mg/L (FEU). A D-dimer result <0.50 mg/L (FEU) was considered negative and a D-dimer result ≥0.50 mg/L (FEU) was considered positive.

Patients with a positive D-dimer result were evaluated by imaging methods, e.g. compression ultrasound and/or venography. Patients with a negative D-dimer, as well as those with negative imaging results, were followed for three months to evaluate potential development of DVT. All patients were subject to imaging at the physician's discretion.

The overall prevalence of DVT in those patients available for final analysis was 21.8 % (93/426). The following instrument-specific sensitivity, specificity and negative predictive value (NPV) with upper and lower 95% confidence limits (CL) were obtained with the INNOVANCE® D-Dimer clinical cutoff of 0.50 mg/L (FEU).

All Patients

Instrument	DVT Patients (n)	Cutoff mg/L FEU	Sensitivity (CL) %	Specificity (CL) %	NPV (CL) %
BCS®/BCS®XP System	426	0.50	100.0 (96.1 – 100.0)	34.5 (29.4 – 39.9)	100.0 (96.8 – 100.0)

Patients with unlikely pre-test probability

Instrument	DVT Patients (n)	Cutoff mg/L FEU	Sensitivity (CL) %	Specificity (CL) %	NPV (CL) %
BCS [®] /BCS [®] XP System	267	0.50	100.0 (83.9 – 100.0)	37.0 (31.0 – 43.4)	100.0 (96.0 – 100.0)

CL = lower and upper 95 % confidence limits



DEPARTMENT OF HEALTH & HUMAN SERVICES

Food and Drug Administration
10903 New Hampshire Avenue
Silver Spring, MD 20993-0002

Siemens Healthcare Diagnostics
c/o Ms. Kathleen Ann Dray-Lyons
Manager, Regulatory Affairs
500 GBC Drive
P.O. Box 6101
Newark, DE 19714-6101

NOV 29 2010

Re: k093626
Trade/Device Name: INNOVANCE[®] D-Dimer
Regulation Number: 21 CFR 864.7320
Regulation Name: Fibrinogen/Fibrin Degradation Products Assay
Regulatory Class: Class II
Product Code: DAP
Dated: October 26, 2010
Received: October 27, 2010

Dear Ms. Dray-Lyons:

We have reviewed your Section 510(k) premarket notification of intent to market the device referenced above and have determined the device is substantially equivalent (for the indications for use stated in the enclosure) to legally marketed predicate devices marketed in interstate commerce prior to May 28, 1976, the enactment date of the Medical Device Amendments, or to devices that have been reclassified in accordance with the provisions of the Federal Food, Drug, and Cosmetic Act (Act) that do not require approval of a premarket approval application (PMA). You may, therefore, market the device, subject to the general controls provisions of the Act. The general controls provisions of the Act include requirements for annual registration, listing of devices, good manufacturing practice, labeling, and prohibitions against misbranding and adulteration. Please note: CDRH does not evaluate information related to contract liability warranties. We remind you, however, that device labeling must be truthful and not misleading.

If your device is classified (see above) into class II (Special Controls), it may be subject to such additional controls. Existing major regulations affecting your device can be found in Title 21, Code of Federal Regulations (CFR), Parts 800 to 895. In addition, FDA may publish further announcements concerning your device in the Federal Register.

Please be advised that FDA's issuance of a substantial equivalence determination does not mean that FDA has made a determination that your device complies with other requirements of the Act or any Federal statutes and regulations administered by other Federal agencies. You must comply with all the Act's requirements, including, but not limited to: registration and listing (21 CFR Part 807); labeling (21 CFR Parts 801 and 809); medical device reporting (reporting of medical device-related adverse events) (21 CFR 803); and good manufacturing practice

requirements as set forth in the quality systems (QS) regulation (21 CFR Part 820). This letter will allow you to begin marketing your device as described in your Section 510(k) premarket notification. The FDA finding of substantial equivalence of your device to a legally marketed predicate device results in a classification for your device and thus, permits your device to proceed to the market.

If you desire specific advice for your device on our labeling regulation (21 CFR Parts 801 and 809), please contact the Office of *In Vitro* Diagnostic Device Evaluation and Safety at (301) 796-5450. Also, please note the regulation entitled, "Misbranding by reference to premarket notification" (21 CFR Part 807.97). For questions regarding the reporting of adverse events under the MDR regulation (21 CFR Part 803), please go to <http://www.fda.gov/MedicalDevices/Safety/ReportaProblem/default.htm> for the CDRH's Office of Surveillance and Biometrics/Division of Postmarket Surveillance.

You may obtain other general information on your responsibilities under the Act from the Division of Small Manufacturers, International and Consumer Assistance at its toll-free number (800) 638-2041 or (301) 796-7100 or at its Internet address <http://www.fda.gov/cdrh/industry/support/index.html>.

Sincerely yours,



Maria M. Chan, Ph.D.
Director
Division of Immunology and Hematology Devices
Office of *In Vitro* Diagnostic Device Evaluation and Safety
Center for Devices and Radiological Health

Enclosure

Indications for Use

510(k) Number (if known):

Device Name: INNOVANCE® D-Dimer

NOV 29 2010

Indications For Use:

INNOVANCE® D-Dimer:

For the quantitative determination of cross-linked fibrin degradation products (D-dimers) in human plasma on Siemens Healthcare Diagnostics and Sysmex® Coagulation Systems. The INNOVANCE® D-Dimer assay is intended for use in conjunction with a non-high clinical pretest probability (PTP) assessment model to exclude deep vein thrombosis (DVT) and pulmonary embolism (PE).

Prescription Use X AND/OR Over-The-Counter-Use
(Per 21 CFR 801 Subpart D) (21 CFR 801 Subpart C)

(PLEASE DO NOT WRITE BELOW THIS LINE – CONTINUE ON ANOTHER PAGE IF NEEDED)

Concurrence of CDRH, Office of Device Evaluation

Page 1 of ____

Maria M Chan
Division Sign-Off

**Office of In Vitro Diagnostic
Device Evaluation and Safety**

510(k) K093626