

**Ultra-TechneKow™ DTE  
(Technetium Tc 99m Generator)**

**Rx Only**

*For the Production of  
Sodium Pertechnetate Tc 99m Injection*

**DESCRIPTION**

The Ultra-TechneKow™ DTE Generator is prepared with fission-produced molybdenum Mo-99 adsorbed onto alumina column shielded by lead, tungsten, or depleted uranium. This generator provides a closed system for the production of sterile metastable technetium Tc-99m, which is produced by the decay of molybdenum Mo-99. Sterile, non-pyrogenic isotonic solutions of Sodium Pertechnetate Tc 99m can be obtained conveniently by periodic aseptic elution of the generator. These solutions should be clear, colorless, and free from any particulate matter.

The carrier-free solution may be used as is, or diluted to the proper concentration. Over the life of the generator, an elution will contain an amount of technetium Tc-99m in direct proportion to the quantity of Mo-99 decay since the previous elution of the generator. The quantity of Tc-99m in the eluate is determined by quantity of Mo-99 on the column and the elapsed time between elutions.

Each eluate of the generator should not contain more than the USP limit of 0.15 kilobecquerel molybdenum Mo-99 per megabecquerel technetium Tc-99m (0.15 microcurie Mo-99 per millicurie Tc-99m) per administered dose at the time of administration and an aluminum ion concentration of not more than 10 micrograms per milliliter of the generator eluate, both of which must be determined by the user before administration.

Since the eluate does not contain an antimicrobial agent, it should not be used after 12 hours from the time of generator elution.

**Physical Characteristics**

Technetium Tc-99m decays by isomeric transition with a physical half-life of 6 hours. The principal photon that is useful for detection and imaging studies is listed in Table 1.

*Table 1. Principal Radiation Emission Data*

Radiation	Mean Percent Per Disintegration	Energy (keV)
-----------	---------------------------------------	-----------------

Gamma-2	89.07	140.5
---------	-------	-------

## External Radiation

The specific gamma ray constant for technetium Tc-99m is 0.795 R/hr-mCi at 1 cm. The first half-value layer is 0.023 cm of lead (Pb). A range of values for the relative attenuation of the radiation emitted by this radionuclide that results from interposition of various thicknesses of Pb is shown in Table 2. For example, the use of 0.27 cm thickness of Pb will attenuate the radiation emitted by a factor of about 1000.

*Table 2. Radiation Attenuation by Lead Shielding*

Shield Thickness (Pb) cm	Coefficient of Attenuation
0.023	0.5
0.09	$10^{-1}$
0.18	$10^{-2}$
0.27	$10^{-3}$

Molybdenum Mo-99 decays to technetium Tc-99m with a molybdenum Mo-99 half-life of 2.75 days, or 66 hours. The physical decay characteristics of molybdenum Mo-99 are such that only 88.6% of the decaying molybdenum Mo-99 atoms form technetium Tc-99m. Generator elutions may be made at any time, but the amount of technetium Tc-99m available will depend on the interval measured from the last elution. Approximately 47% of the maximum available technetium Tc-99m is reached after 6 hours and 95% after 23 hours. To correct for physical decay of molybdenum Mo-99 and technetium Tc-99m, the fractions that remain at selected intervals of time are shown in Tables 3 and 4.

*Table 3. Physical Decay Chart, Molybdenum Mo-99, Half-Life 66 Hours*

Days	Percent Remaining	Days	Percent Remaining
0	100	10	8
1	78	11	6
2	60	12	5
3	47	13	4
4	37	14	3
5	28	15	2
6	22	20	0.6
7	17	25	0.2
8	13	30	0.05
9	10		

*Table 4. Physical Decay Chart; Technetium Tc-99m, Half-Life 6 Hours*

Hours	Percent Remaining	Hours	Percent Remaining
		6	50
		7	45
		8	40
		9	35
		10	32
0*	100	11	28
1	89	12	25
2	79	14	20
3	71	16	16
4	63	18	13
5	56	24	6

\* Calibration Time

## CLINICAL PHARMACOLOGY

The pertechnetate ion distributes in the body similarly to the iodide ion but is not organified when trapped in the thyroid gland. Pertechnetate concentrates in the thyroid gland, salivary glands, stomach and choroid plexus. After intravenous administration it gradually equilibrates with the extracellular space. A fraction is promptly excreted via the kidneys.

Following the administration of Sodium Pertechnetate Tc 99m as an eye drop, the drug mixes with tears within the conjunctival space. Within seconds to minutes it leaves the conjunctival space and escapes into the inferior meatus of the nose through the nasolacrimal drainage system. During this process the pertechnetate ion passes through the canaliculi, the lacrimal sac and the nasolacrimal duct. In the event of any anatomical or functional blockage of the drainage system there will be a backflow resulting in tearing (epiphora). Thus the pertechnetate escapes the conjunctival space in the tears.

While the major part of the pertechnetate escapes within a few minutes of normal drainage and tearing, it has been documented that there is some degree of transconjunctival absorption with turnover of 1.5% per minute in normal individuals, 2.1% per minute in patients without any sac and 2.7% per minute in patients with inflamed conjunctiva due to chronic dacryocystitis. Individual values may vary but these rates are probably representative and indicate that the maximum possible pertechnetate absorbed will remain below one thousandth of that used in other routine diagnostic procedures.

## INDICATIONS AND USAGE

The Ultra-TechneKow™ DTE generator is a source of sodium pertechnetate Tc 99m for use in the preparation of FDA-approved diagnostic radiopharmaceuticals, as described in the labeling of these diagnostic radiopharmaceutical kits.

Sodium Pertechnetate Tc 99m is used **IN ADULTS** as an agent for:

- Thyroid Imaging
- Salivary Gland Imaging
- Urinary Bladder Imaging (direct isotopic cystography) for detection of vesico-ureteral reflux
- Nasolacrimal Drainage System Imaging (dacryoscintigraphy)

Sodium Pertechnetate Tc 99m is used **IN PEDIATRIC PATIENTS** as an agent for:

- Thyroid Imaging
- Urinary Bladder Imaging (direct isotopic cystography) for the detection of vesico-ureteral reflux

## CONTRAINDICATIONS

None.

## WARNINGS

Radiation risks associated with the use of Sodium Pertechnetate Tc 99m are greater in pediatric patients than in adults and, in general, the younger the patient the greater the risk owing to greater absorbed radiation doses and longer life expectancy. These greater risks should be taken firmly into account in all benefit risk assessments involving pediatric patients.

Long-term cumulative radiation exposure may be associated with an increased risk of cancer.

Only use generator eluant specified for use with the Ultra-TechneKow™ DTE Generator. Do not use any other generator eluant or saline from any other source.

## PRECAUTIONS

As in the use of any radioactive material, care should be taken to minimize radiation exposure to the patient consistent with proper patient management and to ensure minimum radiation exposure to occupational workers.

Radiopharmaceuticals should be used only by physicians who are qualified by training and experience in the safe use and handling of radionuclides and whose experience and

training have been approved by the appropriate government agency authorized to license the use of radionuclides.

After the termination of the nasolacrimal imaging procedure, blowing the nose and washing the eyes with sterile distilled water or an isotonic sodium chloride solution will further minimize the radiation dose.

Since the eluate does not contain an antimicrobial agent, it should not be used after 12 hours from time of generator elution.

### **Carcinogenesis, Mutagenesis, Impairment of Fertility**

No long-term animal studies have been performed to evaluate carcinogenic or mutagenic potential or whether Sodium Pertechnetate Tc 99m may affect fertility in males or females.

### **Pregnancy Category C**

Animal reproductive studies have not been conducted with Sodium Pertechnetate Tc 99m. It is also not known whether Sodium Pertechnetate Tc 99m can cause fetal harm when administered to a pregnant woman or can affect reproductive capacity. Sodium Pertechnetate Tc 99m should be given to pregnant women only if the expected benefits to be gained clearly outweigh the potential hazards.

Ideally, examinations using radiopharmaceutical drug products - especially those elective in nature - of women of childbearing capability should be performed during the first ten days following the onset of menses.

### **Nursing Mothers**

Technetium Tc-99m is excreted in human milk during lactation, therefore, formula-feedings should be substituted for breast-feedings.

### **Pediatric Use**

See **INDICATIONS AND USAGE** and **DOSAGE AND ADMINISTRATION** sections. Also see the description of additional risk under **WARNINGS**.

## **ADVERSE REACTIONS**

Allergic reactions including anaphylaxis have been reported infrequently following the administration of Sodium Pertechnetate Tc 99m.

## **DOSAGE AND ADMINISTRATION**

Sodium Pertechnetate Tc 99m is administered by intravenous injection. When imaging the nasolacrimal drainage system, instill the Sodium Pertechnetate Tc 99m by the use of a device such as a micropipette or similar method which will ensure the accuracy of the

dose.

For imaging the urinary bladder and ureters (direct isotopic cystography), the Sodium Pertechnetate Tc 99m is administered by direct instillation aseptically into the bladder via a urethral catheter, following which the catheter is flushed with approximately 200 mL of sterile saline directly into the bladder.

The suggested dose ranges employed for various diagnostic indications in the average ADULT PATIENT (70 kg) are as follows:

Vesico-ureteral imaging: 18.5 to 37 MBq (0.5 to 1 mCi)  
Thyroid gland imaging: 37 to 370 MBq (1 to 10 mCi)  
Salivary gland imaging: 37 to 185 MBq (1 to 5 mCi)  
Nasolacrimal drainage system: Maximum dose of 3.7 MBq (100 µCi)

The recommended dosages in PEDIATRIC PATIENTS are:

Vesico-ureteral imaging: 18.5 to 37 MBq (0.5 to 1 mCi)  
Thyroid gland imaging: 2.22 to 2.96 MBq (60 to 80 µCi) per kg body weight

The patient dose should be measured by a suitable radioactivity calibration system immediately prior to administration.

Parenteral drug products should be inspected visually for particulate matter and discoloration prior to administration whenever solution and container permit. The solution to be administered as the patient dose should be clear, colorless, and contain no particulate matter.

### Radiation Dosimetry

The estimated absorbed radiation doses to an average **ADULT** and **PEDIATRIC** patient from an intravenous injection of various doses of Sodium Pertechnetate Tc 99m distributed uniformly in the total body are shown in tables 5 and 6.

*Table 5. Adult absorbed Radiation Doses from Intravenous Injection*

Organ	Absorbed Radiation Dose (mGy) for a 1110 MBq (30mCi) dose
Adrenals	4.1
Urinary Bladder Wall	20
Bone Surfaces	6.2
Brain	2.2
Breasts	2
Gallbladder Wall	8.3
Stomach Wall	29

Small Intestine	18
ULI Wall	63
LLI Wall	23
Heart Wall	3.5
Kidneys	6
Liver	4.7
Lungs	2.9
Muscle	3.6
Ovaries	11
Pancreas	6.3
Red Marrow	4.1
Skin	2
Spleen	4.8
Testes	3.1
Thymus	2.7
Thyroid	24
Uterus	9
Remaining Tissues	3.9
<b>Effective Dose (mSv)</b>	<b>14</b>

To obtain radiation absorbed dose in rads (30 mCi dose) from the above table, divide individual organ values by a factor of 10 (does not apply for effective dose).

Table 6. Pediatric Absorbed Radiation Doses (mGy) from Intravenous Injection

Age	15 years	10 years	5 years	1 year
<b>Administered activity in MBq (mCi)</b>	<b>1110 (30)</b>	<b>740 (20)</b>	<b>555 (15)</b>	<b>370 (10)</b>
<b>Organ</b>				
Adrenals	5.3	5.4	6.2	7.1
Urinary Bladder Wall	26	22	18	22
Bone Surfaces	7.6	7.5	8.1	10
Brain	2.8	3.1	3.7	4.5
Breasts	2.6	2.6	3.2	4.1
Gallbladder Wall	11	12	13	13
Stomach Wall	38	36	43	59
Small Intestine	22	23	26	30
ULI Wall	81	89	110	140
LLI Wall	31	33	40	48

Heart Wall	4.5	4.6	5.2	6.4
Kidneys	7.2	6.9	7.8	8.5
Liver	6	6.7	8	9.1
Lungs	3.8	3.8	4.4	5.3
Muscle	4.5	4.5	5	6
Ovaries	14	13	14	17
Pancreas	8.1	8.2	8.9	10
Red Marrow	5.1	5	5.2	6
Skin	2.5	2.6	3.2	3.8
Spleen	6	6	6.7	7.8
Testes	4.1	4.3	4.9	6
Thymus	3.6	3.5	4.2	5.3
Thyroid	40	41	67	81
Uterus	11	11	12	14
Remaining Tissues	4.8	4.8	5.4	6.4
<b>Effective Dose (mSv)</b>	19	19	23	29

To obtain radiation absorbed dose in rads (30 mCi dose) from the above table, divide individual organ values by a factor of 10 (does not apply for effective dose).

The estimated absorbed radiation doses to an ADULT patient from the nasolacrimal imaging procedure using a maximum dose of 3.7 megabecquerels (100 microcuries) of Sodium Pertechnetate Tc 99m are shown in Table 7.

*Table 7. Absorbed Radiation Doses from Dacryoscintigraphy*

Tissue	3.7 MBq (100 µCi) Dose of Sodium Pertechnetate Tc 99m	
	mGy	rad
Eye Lens:		
If lacrimal fluid turnover is 16%/min	0.140	0.014
If lacrimal fluid turnover is 100%/min	0.022	0.002
If drainage system is blocked	4.020	0.402
Total Body*	0.011	0.001
Ovaries*	0.030	0.003
Testes*	0.009	0.001
Thyroid*	0.130	0.013

\* Assuming no blockage of draining system

In pediatric patients, an average 30 minute exposure to 37 MBq (1 mCi) of Tc-99m pertechnetate following instillation for direct cystography, will result in the following



estimated radiation doses:

*Table 8. Absorbed Radiation Doses from Cystography (PEDIATRIC)*

<b>Age</b>	<b>Bladder wall dose, mGy (rad)</b>	<b>Gonadal dose, mGy (rad)</b>
1 year	3.6 (0.36)	0.15 (0.015)
5 years	2.0 (0.2)	0.095 (0.0095)
10 years	1.3 (0.13)	0.066 (0.0066)
15 years	0.92 (0.092)	0.046 (0.0046)

## HOW SUPPLIED

The Ultra-TechneKow™ DTE (Technetium Tc 99m) Generators contain the following amount of molybdenum Mo-99 at the date and time of calibration stated on the label.

Catalog No.

883 NDC 0019-9883-03	37 gigabecquerels	(1.0 curie)
884 NDC 0019-9884-04	55.5 gigabecquerels	(1.5 curies)
885 NDC 0019-9885-05	74 gigabecquerels	(2.0 curies)
886 NDC 0019-9886-06	92.5 gigabecquerels	(2.5 curies)
887 NDC 0019-9887-07	111 gigabecquerels	(3.0 curies)
888 NDC 0019-9888-08	129.5 gigabecquerels	(3.5 curies)
889 NDC 0019-9889-09	185 gigabecquerels	(5.0 curies)
890 NDC 0019-9890-10	222 gigabecquerels	(6.0 curies)
891 NDC 0019-9891-11	277.5 gigabecquerels	(7.5 curies)
892 NDC 0019-9892-12	407 gigabecquerels	(11.0 curies)
893 NDC 0019-9893-13	518 gigabecquerels	(14.0 curies)
894 NDC 0019-9894-14	592 gigabecquerels	(16.0 curies)

Each generator is supplied with the following components for the elution of the generator:

- 7 - Evacuated Collecting Vials, 10 mL, Sterile, Non-Pyrogenic
- or
- 5 - Evacuated Collecting Vials, 20 mL, Sterile, Non-Pyrogenic
- 7 - 70% (v/v) Isopropyl Alcohol Wipes
- 7 - Pressure-sensitive "Caution - Radioactive Material" collecting vial labels
- 7 - Pressure-sensitive radioassay data labels for lead elution shield
- 1 - Generator Eluant Vial, 135 mL, Sterile, Non-Pyrogenic
- or
- 2 - Generator Eluant Vials, 135 mL, Sterile, Non-Pyrogenic
- 1 - TechneStat™ Vial, 5 mL, containing 0.5 mL of 1.5 mg/mL methylparaben and 0.2 mg/mL propylparaben, Sterile, Non-pyrogenic
- 1 - Package Insert

The sterile, non-pyrogenic solution used to elute the generator column contains 0.9% sodium chloride. The eluant does not contain an antimicrobial agent.

**EVACUATED COLLECTING VIALS.** Collecting vials are available on request in 10 and 20 milliliter sizes.

## Storage

Store generator and Sodium Pertechnetate Tc 99m solution at controlled room temperature 20° to 25°C (68° to 77°F) [see USP Controlled Room Temperature].

### Expiration Date

The generator should not be used after the expiration date stated on the label.

The expiration time of the Sodium Pertechnetate Tc 99m solution is not later than 12 hours after time of elution. If the eluate is used to reconstitute a kit, the radiolabeled kit should not be used after 12 hours from the time of generator elution or after the expiration time stated on the labeling for the prepared drug, whichever is earlier.

## Directions for Use of the Technetium Tc 99m Generator

- NOTE 1: Immediately upon delivery, the generator should be placed within a minimum of one-inch of lead shielding in such a manner so as to minimize radiation exposure to attending personnel.
- NOTE 2: Wear waterproof gloves during the elution procedure and during subsequent reconstitution of kits with the eluate.
- NOTE 3: Use a shielded syringe to withdraw patient dose or to transfer Sodium Pertechnetate Tc 99m into mixing vials during kit reconstitution.

NOTE 4: The needles in the generator are sterile beneath their covers, and the generator has been cleaned underneath the top cover. Additional disinfection of these areas with agents containing alcohol may unfavorably influence the Tc 99m yield.

Eluting the generator every 24 hours will provide optimal amounts of Sodium Pertechnetate Tc 99m. However, the generator may be eluted whenever sufficient amounts of technetium Tc-99m have accumulated within the column.

#### For Example

<u>Time After First Elution (hrs.)</u>	<u>Approximate Yield (% of First Elution)</u>
1	10
2	19
3	27
4	35
5	41
6	47

#### Preparation

**Note:** The following instructions are applicable for operation of the Ultra-TechneKow™ DTE Generator with or without the utilization of the alignment adaptor.

1. Rotate the top cover 30° counterclockwise and lift up to remove.
2. Lift the generator by its handle and position it inside the auxiliary shield, aligning the notch in the elution station with the front of the auxiliary shield. Move the handle so that it is not covering the generator top by pushing it off to the side in between the generator and the auxiliary shield.
3. Remove the flip-top cap of the eluant vial; disinfect the stopper, allowing the stopper to dry before use. Remove and store the needle cover from the eluant needles; invert the eluant vial and push down into place on the eluant needles.
4. Place the alignment adaptor onto the top of the generator with the raised portion of the adaptor located over the elution station.
5. Remove the flip-top cap of the TechneStat™ vial; disinfect the stopper, allowing the stopper to dry before use. Secure the TechneStat™ vial into the TechneStat™ vial shield.
6. Remove and store the needle cover from the elution needle. Place the auxiliary shield lid onto the top auxiliary shield ring ensuring alignment of the key-hole with the elution needle.
7. Carefully lower the TechneStat™ vial shield containing the TechneStat™ vial through the key-hole, inserting the shielded TechneStat™ vial onto the elution needle.

#### Elution

1. Remove the flip-top cap of the appropriate evacuated vial; disinfect the stopper, allowing the stopper to dry before use. Place the evacuated vial into the elution shield utilizing the spacer if required.

2. Remove the shielded TechneStat™ vial by carefully lifting the TechneStat™ vial shield from the elution needle. Position the shielded evacuated vial by carefully lowering the elution shield into the elution station. Piercing the septum of the evacuated vial with the elution needle will begin the elution process.
3. Wait until the evacuated vial has completely filled itself. Depending on the size of the evacuated vial, this may take a few minutes. **Never interrupt the elution by lifting the elution shield! NOTE: Do not use generator eluate if its appearance is discolored.**
4. Carefully remove the elution shield and replace with the shielded TechneStat™ vial (see Step 6 of the **Preparation** section).
5. Determine the technetium Tc-99m concentration and molybdenum Mo-99 content for dispensing purposes. The generator eluate may be assayed using an appropriate detection system. The manufacturer's instructions for operation of the instrument/equipment should be followed for measurement of Technetium Tc-99m and Molybdenum Mo-99 activity. **NOTE: Molybdenum Mo-99 Breakthrough Limit** – The acceptable limit is 0.15 kilobecquerel molybdenum Mo-99 per megabecquerel technetium Tc-99m (0.15 microcurie Mo-99 per millicurie Tc-99m) per administered dose in the Injection, at the time of administration (see USP, Sodium Pertechnetate Tc 99m Injection).
6. Determine the aluminum ion concentration of the eluate. **NOTE: Aluminum Ion Breakthrough Limit** - The acceptable limit is not more than 10 micrograms per milliliter of eluate (see USP, Sodium Pertechnetate Tc 99m Injection).

### Subsequent Elutions

Repeat steps 1 through 6 of the Elution procedure above.

### Vacuum Loss

If the vacuum in the collecting vial is lost, do not attempt to re-evacuate the vial, but discard and use a new collecting vial.

### Expired Generator Disposal

1. Following the life of the generator, remove and dispose of the used TechneStat™ vial and the eluant vial.
2. If appropriate, remove and store the Alignment Adaptor for use with replacement generator.
3. Cover the elution and eluant needles with the stored needle covers.
4. Close the generator system with its top cover by rotating with downward pressure.
5. The intact generator assembly should be either returned to Mallinckrodt Inc. or disposed of in accordance with applicable regulations.

This generator is approved for use by persons licensed by the U.S. Nuclear Regulatory Commission to use by-product material identified in Section 35.200 or under equivalent licenses of Agreement States.

Ultra-TechneKow and TechneStat are trademarks of Mallinckrodt LLC.

Manufactured by:  
Mallinckrodt Inc.  
St. Louis, MO 63134 USA

Distributed in Canada by:  
Mallinckrodt Canada ULC  
Kirkland, QC, Canada, H9J 3Z4  
Estb. Lic.: 102017-B

A880I0

Rev 01/2014

**Mallinckrodt**