

HIGHLIGHTS OF PRESCRIBING INFORMATION

These highlights do not include all the information needed to use **CARDIOGEN-82** safely and effectively. See full prescribing information for **CARDIOGEN-82**.

CARDIOGEN-82[®] (rubidium Rb 82 generator) to produce rubidium chloride Rb 82 injection, for intravenous use
Initial U.S. Approval: 1989

WARNING: HIGH LEVEL RADIATION EXPOSURE WITH USE OF INCORRECT ELUENT and FAILURE TO FOLLOW THE ELUATE TESTING PROTOCOL

See full prescribing information for complete boxed warning.

High Level Radiation Exposure with Use of Incorrect Eluent

Using the incorrect eluent can cause high Strontium (Sr) 82 and Sr 85 breakthrough levels.

- Use only additive-free 0.9% Sodium Chloride Injection to elute CardioGen-82.
- If an incorrect solution is used to elute CardioGen-82:
 - Immediately stop the patient infusion;
 - Evaluate the patient's radiation absorbed dose and monitor for the effects of radiation to critical organs such as bone marrow;
 - Permanently discontinue use of the affected generator. (2.5, 2.9, 5.1)

Excess Radiation Exposure with Failure to Follow the Eluate Testing Protocol

Excess radiation exposure occurs when the levels of Sr 82 or Sr 85 in the Rubidium Chloride Rb 82 Injection exceed limits.

- Strictly adhere to the generator eluate testing protocol.
- Stop using the generator if it reaches any of its Expiration Limits. (2.5, 2.6, 2.7, 5.2)

RECENT MAJOR CHANGES

Dosage and Administration, 5/2026
Dosing Using the Infusion System Model 510 (2.2)-Removed
Eluate Testing Protocol for Infusion System Model 510 (2.6)-Removed
Expiration Limits of CardioGen-82 (2.7)

INDICATIONS AND USAGE

CardioGen-82, used to produce Rubidium Chloride Rb 82 Injection, is a radioactive diagnostic agent indicated for positron emission tomography (PET) of the myocardium under rest or pharmacologic stress conditions to evaluate regional myocardial perfusion in adult patients with suspected or existing coronary artery disease. (1)

FULL PRESCRIBING INFORMATION: CONTENTS*

WARNING: HIGH LEVEL RADIATION EXPOSURE WITH USE OF INCORRECT ELUENT and FAILURE TO FOLLOW THE ELUATE TESTING PROTOCOL

1 INDICATIONS AND USAGE

2 DOSAGE AND ADMINISTRATION

- 2.1 Radiation Safety – Drug Handling
- 2.2 Dosing and Administration of Rubidium Chloride Rb 82 Injection
- 2.3 Image Acquisition Instructions
- 2.4 Infusion System
- 2.5 Directions for Eluting Rubidium Chloride Rb 82 Injection
- 2.6 Eluate Testing Protocol for Infusion System Model 1700
- 2.7 Expiration Limits of CardioGen-82
- 2.8 Dose Delivery Limit of CardioGen-82
- 2.9 Radiation Dosimetry

3 DOSAGE FORMS AND STRENGTHS

4 CONTRAINDICATIONS

5 WARNINGS AND PRECAUTIONS

- 5.1 High Level Radiation Exposure with Use of Incorrect Eluent
- 5.2 Excess Radiation Exposure with Failure to Follow the Eluate Testing Protocol
- 5.3 Risks Associated with Pharmacologic Stress
- 5.4 Cumulative Radiation Exposure: Long-Term Risk of Cancer

6 ADVERSE REACTIONS

DOSAGE AND ADMINISTRATION

- Dosing: 10 MBq/kg to 30 MBq/kg actual body weight (0.27 mCi/kg to 0.81 mCi/kg) per rest or stress component of a procedure via intravenous infusion at 50 mL/minute or 20 mL/minute. (2.2)
- Do not exceed a maximum dose of 2,220 MBq (60 mCi) or a maximum volume of 100 mL per rest or stress component of a procedure. (2.2)
- The minimum interval between the rest and stress doses is 10 minutes to allow sufficient Rb 82 decay. (2.2)
- Start image acquisition 60 seconds to 90 seconds after completion of the infusion; if a longer circulation time is anticipated, wait for 120 seconds. Image acquisition is 5 minutes long. (2.3)
- For radiation safety, infusion system, elution instruction, eluate testing, dose delivery, and expiration limits of CardioGen-82, and radiation dosimetry see full prescribing information. (2.1, 2.5, 2.6, 2.7, 2.8, 2.9)

DOSAGE FORMS AND STRENGTHS

Rubidium Rb 82 generator used to produce Rubidium Chloride Rb 82 Injection: 3,330 MBq to 5,550 MBq (90 mCi to 150 mCi) of strontium-82 (Sr 82) at calibration time, adsorbed on a hydrous stannic oxide column. (3)

CONTRAINDICATIONS

None. (4)

WARNINGS AND PRECAUTIONS

Risk Associated with Pharmacological Stress: Pharmacologic induction of cardiovascular stress may cause serious adverse reactions such as myocardial infarction, arrhythmia, hypotension, broncho-constriction, and cerebrovascular events. Perform testing only in setting where cardiac resuscitation equipment and trained staff are readily available. (5.3)

ADVERSE REACTIONS

To report SUSPECTED ADVERSE REACTIONS, contact Bracco Diagnostics Inc. at 1-800-257-8151 or FDA at 1-800-FDA-1088 or www.fda.gov/medwatch.

USE IN SPECIFIC POPULATIONS

Lactation: Do not resume breastfeeding until at least one hour after administration of Rubidium Chloride Rb 82 Injection. (8.2)

See 17 for PATIENT COUNSELING INFORMATION.

Revised: 5/2026

8 USE IN SPECIFIC POPULATIONS

- 8.1 Pregnancy
- 8.2 Lactation
- 8.4 Pediatric Use
- 8.5 Geriatric Use

11 DESCRIPTION

- 11.1 Generator Characteristics
- 11.2 Nuclear Physical Characteristics

12 CLINICAL PHARMACOLOGY

- 12.1 Mechanism of Action
- 12.2 Pharmacodynamics
- 12.3 Pharmacokinetics

13 NONCLINICAL TOXICOLOGY

- 13.1 Carcinogenesis, Mutagenesis, Impairment of Fertility

14 CLINICAL STUDIES

16 HOW SUPPLIED/STORAGE AND HANDLING

17 PATIENT COUNSELING INFORMATION

* Sections or subsections omitted from the full prescribing information are not listed.

FULL PRESCRIBING INFORMATION

WARNING: HIGH LEVEL RADIATION EXPOSURE WITH USE OF INCORRECT ELUENT and FAILURE TO FOLLOW THE ELUATE TESTING PROTOCOL

High Level Radiation Exposure with Use of Incorrect Eluent

Patients are exposed to high radiation levels when the CardioGen-82 generator is eluted with the incorrect eluent due to high strontium (Sr) 82 and Sr 85 breakthrough levels.

- Use only additive-free 0.9% Sodium Chloride Injection to elute the generator.
- If an incorrect solution is used to elute the CardioGen-82 generator:
 - Immediately stop the patient infusion;
 - Evaluate the patient's radiation absorbed dose, and monitor for the effects of radiation to critical organs such as bone marrow;
 - Permanently discontinue the use of the affected generator [see *Dosage and Administration (2.5)* and *Warnings and Precautions (5.1)*].

Excess Radiation Exposure with Failure to Follow the Eluate Testing Protocol

Excess radiation exposure occurs when the levels of Sr 82 or Sr 85 in the Rubidium Chloride Rb 82 Injection exceed specified limits.

- Strictly adhere to the generator eluate testing protocol, to minimize the risk of excess radiation exposure, including daily testing and additional testing at Alert Limits.
- Stop using the generator if it reaches any of its Expiration Limits:
 - 17 L for the generator's cumulative eluate volume
 - 60 days post generator calibration date
 - An eluate Sr 82 level of 0.01 microCi /mCi Rb 82
 - An eluate Sr 85 level of 0.1 microCi /mCi Rb 82 [see *Dosage and Administration (2.5, 2.6, 2.7, 2.8)* and *Warnings and Precautions (5.2)*]

1 INDICATIONS AND USAGE

CardioGen-82, used to produce Rubidium Chloride Rb 82 Injection, is indicated for positron emission tomography (PET) of the myocardium under rest or pharmacologic stress conditions to evaluate regional myocardial perfusion in adult patients with suspected or existing coronary artery disease.

2 DOSAGE AND ADMINISTRATION

2.1 Radiation Safety - Drug Handling

CardioGen-82, when eluted with additive-free 0.9% Sodium Chloride Injection, produces Rubidium Chloride Rb 82 Injection. Handle CardioGen-82, Rubidium Chloride Rb 82 Injection, and Infusion Systems with appropriate safety measures to minimize radiation exposure. Wear waterproof gloves and effective shielding throughout the entire preparation and handling [See *Warnings and Precautions (5.4)*].

Limit the use of radiopharmaceuticals to healthcare providers who are qualified by specific training and experience in the safe use and handling of radionuclides and whose experience and training have been approved by the appropriate government agency authorized to license the use of radionuclides.

2.2 Dosing and Administration of Rubidium Chloride Rb 82 Injection

Dosing

The recommended dose of Rubidium Chloride Rb 82 Injection per rest or stress component of a PET MPI procedure in adults is 10 MBq/kg to 30 MBq/kg actual body weight (0.27 mCi/kg to 0.81 mCi/kg) administered by intravenous infusion at 50 mL/minute or 20 mL/minute through a catheter inserted into a large peripheral vein.

Do not exceed a maximum dose of 2,220 MBq (60 mCi) or a maximum volume of 100 mL per rest or stress component of a procedure.

Administration Instructions

- Observe aseptic techniques in all drug handling and administration.
- Use CardioGen-82 with the CardioGen-82 Model 1700 Infusion System to elute and administer Rubidium Chloride Rb 82 Injection.
- Use the lowest dose necessary, consistent with the goal of as low as reasonably achievable (ALARA), to obtain adequate cardiac visualization based on patient body weight and the imaging equipment and acquisition methodology used to perform the procedure. For example, 3D image acquisition may require doses at the lower end of the recommended range, compared to 2D imaging.

- Administer two single doses to complete rest and stress myocardial perfusion imaging (MPI) [see *Dosage and Administration (2.3)*].
- The minimum interval between the rest and stress doses is 10 minutes to allow sufficient Rb 82 decay.
- Instruct patients to void as soon as a study is completed and as often as possible thereafter for at least one hour [see *Warnings and Precautions (5.4)*].

2.3 Image Acquisition Instructions

Rest Imaging

- Administer a single dose of Rubidium Chloride Rb 82 Injection.
- Start imaging 60 seconds to 90 seconds after completion of the first dose infusion and acquire images for 5 minutes.

Stress Imaging

- Begin the study at least 10 minutes after completion of the resting dose infusion, to allow for sufficient Rb 82 decay.
- Administer a pharmacologic stress agent in accordance with its prescribing information.
- After the administration of the pharmacologic stress agent, administer the second dose of Rubidium Chloride Rb 82 Injection at the time interval according to the prescribing information of the pharmacologic stress agent.
- Start imaging 60 seconds to 90 seconds after completion of the second dose infusion and acquire images for 5 minutes.

Rest and Stress Imaging

- If a longer circulation time is anticipated (e.g., in a patient with severe left ventricular dysfunction), start imaging 120 seconds after the dose.
- Acquisition may be started immediately post-injection if dynamic imaging is needed.

2.4 Infusion System

- Use CardioGen-82 only with the CardioGen-82 Infusion System Model 1700, which is specifically designed for use with the generator and capable of accurate measurement and delivery of doses of Rubidium Chloride Rb 82 Injection.
- The accuracy of the CardioGen-82 Infusion System Model 1700 for measuring Rubidium Rb 82 Chloride Injection is as follows:
 - For doses of 1,110 MBq to 2,220 MBq (30 mCi to 60 mCi), the system provides an accuracy of $\pm 10\%$.
 - For doses of 370 MBq to 1,110 MBq (10 mCi to 30 mCi), the system provides an accuracy of $\pm 15\%$.
- Follow instructions in the CardioGen-82 Infusion System Model 1700 Operator's Manual for the setup and administration of Rubidium Chloride Rb 82 Injection.

2.5 Directions for Eluting Rubidium Chloride Rb 82 Injection

- **Use only additive-free 0.9% Sodium Chloride Injection to elute the generator** [see *Warnings and Precautions (5.1)*].
- Prepare the 0.9% Sodium Chloride Injection for use with the Saline Tag.
 - Prepare the intravenous port in accordance with the approved prescribing information of the 0.9% Sodium Chloride Injection.
 - The intravenous administration port of the 0.9% Sodium Chloride Injection container must be penetrated only one time.
 - Strap the saline tag provided with the CardioGen-82 Infusion System on the additive-free 0.9% Sodium Chloride Injection container and install on the CardioGen-82 Infusion System.
 - Once the container port closure is penetrated, it should remain installed on the CardioGen-82 Infusion System for its entire period of use. A maximum of 12 hours from the initial port closure penetration is permitted, after which the bag must be replaced for the next patient.
- Allow at least 10 minutes between elutions for regeneration of Rb 82.
- The system automatically discards the first 50 mL of eluate each day the generator is eluted and records and saves all eluate volumes (all flushing, quality control (QC) testing, and patient infusions), representing the cumulative volume of eluate each day the generator is eluted.
- Perform eluate testing according to the Eluate Testing Protocol [see *Dosage and Administration (2.6)*].
- Stop using the generator when the expiration limits are reached [see *Dosage and Administration (2.7)*].
- The maximum available activity of Rb 82 (dose delivery limit) will decrease as the generator ages. See Table 1 for the estimated maximum available Rb 82 activity as a function of generator age for a generator with an initial 100 mCi of Sr 82 [see *Dosage and Administration (2.8)*].

2.6 Eluate Testing Protocol for Infusion System Model 1700

General Precautions

- **Use only additive-free 0.9% Sodium Chloride Injection for all elutions** [see *Warnings and Precautions (5.1)*].
- Observe aseptic technique throughout.
- Follow all instructions in the CardioGen-82 Infusion System Model 1700 Operator's Manual for performing all eluate testing as described.

Mandatory Eluate Testing

Perform Mandatory Eluate Testing to determine Rb 82, Sr 82, and Sr 85 levels (see *Daily Quality Control Procedure*):

- Daily – Before administering Rubidium Chloride Rb 82 Injection to the first patient each day.
- Repeat as indicated after an Alert Limit has been detected.

Alert Limits and System Response:

The CardioGen-82 Infusion System Model 1700 will automatically indicate when any of the following Alert Limits have been reached:

- 14 L total elution volume has passed through the generator column, or
- Sr 82 level reaches 0.002 microCi per mCi Rb 82, or
- Sr 85 level reaches 0.02 microCi per mCi Rb 82.

Once an Alert Limit has been reached, the system will require additional eluate testing at intervals determined by the day's elution volume and enforced by the system software. Specifically, the infusion system will require the user to perform additional eluate testing after each 750 mL of elution volume when any Alert Limit parameter has been reached.

Infusion System Calibration

Before administering Rubidium Chloride Rb 82 Injection to the first patient, perform the Rubidium Chloride Rb 82 Injection Dose Calibration using an external dose calibrator at the following times (see *Calibration Procedure*):

- After installation of a new generator
- After installation of a new CardioGen-82 Accessory Package, Item # 001710
- Every 14 days

Calibration Procedure

1. Set a dose calibrator for Rb 82 as recommended by the manufacturer. Obtain the reading from the instrument in millicuries.
2. Following the prompts in the Graphical User Interface (GUI) for the CardioGen-82 Infusion System Model 1700, elute the generator with additive-free 0.9% Sodium Chloride Injection at a rate of 50 mL/min and collect the eluate in the stoppered vial specifically provided for use with the CardioGen-82 Infusion System Model 1700 (alternative vials, glass or plastic are not suitable). Note the exact time of end of elution (EOE).
3. Using the external dose calibrator, assay the eluate at exactly 2:30, 3:45, or 5:00 minutes after EOE.
4. Following the prompts in the GUI for the CardioGen-82 Infusion System Model 1700, enter the Rb 82 reading from the dose calibrator and the time since EOE.
5. The infusion system software will automatically calculate the Calibration Ratio.
 - If the ratio is within +/- 2% (0.98 to 1.02), the infusion system will allow acceptance of the calibration factor that was used for the elution.
 - If the ratio is not within +/- 2% (0.98 to 1.02), the system requires another calibration elution (steps 1 through 4).
6. Repeat steps 1 through 4 for a flow rate of 20 mL/min.

Daily Quality Control Procedure

Each day, before administering Rubidium Chloride Rb 82 Injection, perform the following testing:

1. Place the stoppered vial into the Sr detector well. Use only the stoppered vial provided for use with the CardioGen-82 Infusion System Model 1700. Following the GUI prompts, initiate the Daily Quality Control workflow.
2. The infusion system will automatically perform the Sr Detector Background Reading.
3. The infusion system will automatically perform the Generator Column Wash.
4. Strontium Level Test and Dose Constancy:
 - a. The infusion system will elute the generator with 50 mL of additive-free 0.9% Sodium Chloride Injection at a rate of 50 mL/min into the stoppered vial.
 - b. The Sr detector measures the Rb 82 and strontium in the 50 mL elution.
 - c. The infusion system software will automatically calculate the Sr 82 and Sr 85 levels on the day (post calibration) of the measurement using the ratio of Sr 85/Sr 82 on the day of calibration provided on the generator label, and using the full exponential decay calculation for each, accounting for the generator's age.

- d. Using the Rb 82 and strontium measurements, the infusion system software will automatically calculate microCi Sr 82/mCi Rb 82 and microCi Sr 85/mCi Rb 82. The GUI will automatically indicate if the results exceed Alert or Expiration Limits.
 - e. The infusion system software will automatically calculate Dose Constancy.
5. Constancy Check of the Sr detector: The GUI will prompt the user to perform the constancy check of the Sr detector.
- a. Place the external constancy source in the detector well of the infusion system.
 - b. The infusion system software will automatically calculate the constancy of the Sr detector versus the external constancy source when instructed.

2.7 Expiration Limits of CardioGen-82

The CardioGen-82 Infusion System Model 1700 software will automatically indicate and will stop use of CardioGen-82 once any one of the following Expiration Limits is reached.

- A total elution volume of 17 L has passed through the generator column
- 60 days post calibration date
- An eluate Sr 82 level of 0.01 microCi/mCi Rb 82
- An eluate Sr 85 level of 0.1 microCi/mCi Rb 82

2.8 Dose Delivery Limit of CardioGen-82

The maximum available Rb 82 activity per elution (dose delivery limit) will decrease as the generator ages. Certain doses, including the maximum recommended dose of 2,220 MBq (60 mCi), may not be achievable for the entire shelf-life of the generator. Table 1 provides an estimate of the maximum available activity of Rb 82 as a function of generator age for a generator with an initial 100 mCi of Sr 82.

Generator Age (days)²	Maximum Rb 82 Activity per Elution
0 to 17	60 mCi (2,220 MBq)
24	50 mCi (1,850 MBq)
32	40 mCi (1,480 MBq)
42	30 mCi (1,110 MBq)
57	20 mCi (740 MBq)
60	19 mCi (703 MBq)

¹Estimate is based on a 100 mCi (3,700 MBq) Sr 82 generator at calibration.

²Generator age at which delivery limit is reached varies with generator activity at release. For example, a 90 mCi (3,330 MBq) generator and a 150 mCi (5,550 MBq) generator will reach a delivery limit <60 mCi at ≥ 14 days and ≥ 33 days, respectively.

2.9 Radiation Dosimetry

The estimated absorbed radiation doses for Rb 82, Sr 82, and Sr 85 from an intravenous injection of Rubidium Chloride Rb 82 Injection are shown in Table 2.

Organ^{a,b}	Rb 82 (Average for Rest and Stress) mrem/mCi (microSv/3.7MBq)^c	Sr 82 mrem/microCi (microSv/3.7kBq)^c	Sr 85 mrem/microCi (microSv/3.7kBq)^c
Adrenals	7.56	10.6	5.03
Bone – Osteogenic cells	1.86	---	---
Bone Surface	---	107	9.81
Brain	0.6	8.29	2.96
Breast	0.82	7.03	1.72
Gall Bladder Wall	3.17	8.47	2.82
Heart Wall	16.5	8.18	2.67
Kidneys	20.04	9.18	2.5
Liver	4.2	8.10	2.5
Lower Large Intestine Wall	2.84	51.8	5.14
Lungs ^d	10.7	8.25	2.84
Muscles	1.29	8.14	2.66

Table 2. Estimated Absorbed Radiation Dose Coefficient in Adults

Organ^{a,b}	Rb 82 (Average for Rest and Stress) mrem/mCi (microSv/3.7MBq)^c	Sr 82 mrem/microCi (microSv/3.7kBq)^c	Sr 85 mrem/microCi (microSv/3.7kBq)^c
Ovaries	1.41	10.2	4.29
Pancreas	8.85	9.1	3.46
Red Marrow	1.19	91	9.84
Skin	1.14	7.03	1.75
Small Intestine	4.76	9.62	4.03
Spleen	6.61	8.1	2.54
Stomach	8.14	7.84	2.26
Testes	0.82	7.25	1.7
Thymus	1.49	7.84	2.33
Thyroid	6.11	8.07	2.57
Upper Large Intestine	5.94	23.7	3.62
Urinary Bladder Wall	1.61	21.9	2.9
Uterus	3.72	9.14	3.32
Total Body	1.77	Not calculated	Not calculated
Effective Dose ^e	4.74 ^f	23.4	4.03

^a Rb 82 doses are averages of rest and stress dosimetry data. To calculate organ doses (mrem) from Rb 82, multiply the dose coefficient for each organ by the administered activity in mCi.

^b Sr 82 and Sr 85 doses are calculated using software package DCAL and ICRP dose coefficients. To calculate organ doses (mrem) attributable to Sr 82, and Sr 85, multiply the dose coefficients by the calculated amounts of strontium in microCi.

^c To convert to SI units, insert the dose coefficient into the formula in parentheses, e.g. for adrenals $7.56 \text{ mrem/mCi} = 7.56 \text{ microSv}/37 \text{ MBq} = 2.04 \times 10^{-13} \text{ Sv/Bq}$.

^d Calculated from ICRP 66

^e Calculated from ICRP 60

^f Stress phase only

3 DOSAGE FORMS AND STRENGTHS

Rubidium Rb 82 generator used to produce Rubidium Chloride Rb 82 Injection: strontium-82 (Sr 82), with an activity of 3,330 MBq to 5,550 MBq (90 mCi to 150 mCi) at calibration time, adsorbed on a hydrous stannic oxide column.

4 CONTRAINDICATIONS

None.

5 WARNINGS AND PRECAUTIONS

5.1 High Level Radiation Exposure with Use of Incorrect Eluent

Additives present in solutions (particularly calcium ions) used by mistake to elute CardioGen-82 expose patients to high levels of radiation by causing the release of large amounts of Sr 82 and Sr 85 into the eluate regardless of the generator's age or prior use. When solutions containing calcium ions are used, high levels of radioactivity are present in any subsequent eluate even with the use of additive-free 0.9% Sodium Chloride Injection.

Use only additive-free 0.9% Sodium Chloride Injection to elute CardioGen-82. If an incorrect eluent is used, immediately stop the patient infusion, evaluate the patient's radiation absorbed dose, and monitor for the effects of radiation to critical organs such as bone marrow. Permanently discontinue use of the affected CardioGen-82 generator [*see Dosage and Administration (2.5)*].

5.2 Excess Radiation Exposure with Failure to Follow Eluate Testing Protocol

Excess radiation exposure occurs when the Sr 82 and Sr 85 levels in the Rubidium Chloride Rb 82 Injection exceed the specified generator eluate limits.

Strictly adhere to the eluate testing protocol to minimize radiation exposure to the patient. Stop using the CardioGen-82 generator when the expiration limits are reached [*see Dosage and Administration (2.6, 2.7, 2.8)*].

5.3 Risk Associated with Pharmacologic Stress

Pharmacologic induction of cardiovascular stress may be associated with serious adverse reactions such as myocardial infarction, arrhythmia, hypotension, bronchoconstriction, and cerebrovascular events. Perform pharmacologic stress testing

in accordance with the pharmacologic stress agent's prescribing information and only in the setting where cardiac resuscitation equipment and trained staff are readily available.

5.4 Cumulative Radiation Exposure: Long-Term Risk of Cancer

Rubidium chloride Rb 82 contributes to a patient's overall long-term cumulative radiation exposure. Long-term cumulative radiation exposure is associated with an increased risk of cancer. Use the lowest dose of Rubidium Chloride Rb 82 Injection necessary for imaging and ensure safe handling to protect the patient and health care providers. Encourage patients to void as soon as a study is completed and as often as possible thereafter for a least one hour [*see Dosage and Administration (2.1, 2.2)*].

6 ADVERSE REACTIONS

The following clinically significant adverse reactions are described elsewhere in the labeling:

- High Level Radiation Exposure with Use of Incorrect Eluent [*see Warnings and Precautions (5.1)*]
- Excess Radiation Exposure with Failure to Follow Eluate Testing Protocol [*see Warnings and Precautions (5.2)*]

The following serious adverse reactions have been identified during post-approval use of CardioGen-82. Because these reactions are reported voluntarily from a population of uncertain size, it is not always possible to reliably estimate their frequency or establish a causal relationship to drug exposure.

Radiation Exposure

High level radiation exposure to the bone marrow from using an incorrect eluent.

Excess radiation exposure due to insufficient eluate testing.

8 USE IN SPECIFIC POPULATIONS

8.1 Pregnancy

Risk Summary

There are no data available on the use of rubidium chloride Rb 82 in pregnant women. Animal reproductive studies have not been conducted with rubidium chloride Rb 82. However, all radiopharmaceuticals, including Rubidium Chloride Rb 82 Injection generated by CardioGen-82, have the potential to cause fetal harm depending on the fetal stage of development and the magnitude of the radiation dose. If considering Rubidium Chloride Rb 82 Injection administration to a pregnant woman, inform the patient about the potential for adverse pregnancy outcomes based on the radiation dose from rubidium-82 (Rb 82) and the gestational timing of exposure.

The estimated background risk of major birth defects and miscarriage for the indicated population is unknown. In the U.S. general population, the estimated background risk of major birth defects and miscarriage in clinically recognized pregnancies is 2% to 4% and 15% to 20%, respectively.

8.2 Lactation

Risk Summary

There is no information regarding the presence of rubidium chloride Rb 82 in human milk, the effects on the breastfed infant or the effects on milk production. Due to the short half-life of Rb 82 (75 seconds), exposure of a breastfed infant through breast milk can be minimized by temporary discontinuation of breastfeeding (*see Clinical Considerations*). The developmental and health benefits of breastfeeding should be considered along with the mother's clinical need for Rubidium Chloride Rb 82 Injection, any potential adverse effects on the breastfed child from Rb 82 or from the underlying maternal condition.

Clinical Considerations

Exposure to Rb 82 through breast milk can be minimized if breastfeeding is discontinued when Rubidium Chloride Rb 82 Injection is administered. Do not resume breastfeeding until at least one hour after administration of Rubidium Chloride Rb 82 Injection.

8.4 Pediatric Use

The safety and effectiveness of CardioGen-82, used to produce Rubidium Chloride Rb 82 Injection, in pediatric patients have not been established.

8.5 Geriatric Use

In elderly patients with a clinically important decrease in cardiac function, lengthen the delay between infusion and image acquisition [see *Dosage and Administration (2.3)*]. Observe for the possibility of fluid overload.

11 DESCRIPTION

11.1 Generator Characteristics

CardioGen-82 (rubidium Rb 82 generator) contains accelerator-produced Sr 82 adsorbed on stannic oxide in a lead-shielded column and provides a means to produce sterile nonpyrogenic Rubidium Chloride Rb 82 Injection, a radioactive diagnostic agent, for intravenous use. The chemical form of Rb 82 is $^{82}\text{RbCl}$.

The amount (mCi) of Rb 82 obtained in each elution will depend on the potency of the generator. When eluted at a rate of 50 mL/minute, each generator eluate at the end of elution should not contain more than 0.02 microCi of Sr 82 and not more than 0.2 microCi of Sr 85 per mCi of Rubidium Chloride Rb 82 Injection, and not more than 1 mcg of tin per mL of eluate.

11.2 Nuclear Physical Characteristics

Rb 82 decays by positron emission and associated gamma emission with a physical half-life of 75 seconds. Table 3 shows the annihilation photons released following positron emission which are useful for detection and imaging studies.

The decay modes of Rb 82 are: 95.5% by positron emission, resulting in the production of annihilation radiation, i.e., two 511 keV gamma rays; and 4.5% by electron capture, resulting in the emission of "prompt" gamma rays of predominantly 776.5 keV. Both decay modes lead directly to the formation of stable krypton 82 (Kr 82).

Table 3. Principal Radiation Emission Data for Rb 82

Radiation	Mean Percent Per Disintegration	Mean Energy (keV)
Annihilation photons (2)	191.01	511 (each)
Gamma rays	13-15	776.5

The specific gamma ray constant for Rb 82 is 6.1 R/hour-mCi at 1 cm. The first half-value layer is 0.7 cm of lead (Pb). Table 4 shows a range of values for the relative attenuation of the radiation emitted by this radionuclide that results from interposition of various thicknesses of lead. For example, the use of a 7.0 cm thickness of Pb will attenuate the radiation emitted by a factor of about 1,000.

Table 4. Radiation Attenuation by Lead Shielding

Shield Thickness (Pb) cm	Attenuation Factor
0.7	0.5
2.3	10^{-1}
4.7	10^{-2}
7.0	10^{-3}
9.3	10^{-4}

Sr 82 (half-life of 25 days) decays to Rb 82. To correct for physical decay of Sr 82, Table 5 shows the fractions that remain at selected intervals after the time of calibration.

Table 5. Physical Decay Chart: Sr 82 half-life 25 days

Days	Fraction Remaining	Days	Fraction Remaining	Days	Fraction Remaining
0*	1	21	0.559	41	0.321
1	0.973	22	0.543	42	0.312
2	0.946	23	0.529	43	0.304
3	0.920	24	0.514	44	0.295
4	0.895	25	0.500	45	0.287
5	0.871	26	0.486	46	0.279
6	0.847	27	0.473	47	0.272
7	0.824	28	0.460	48	0.264

8	0.801	29	0.448	49	0.257
9	0.779	30	0.435	50	0.250
10	0.758	31	0.423	51	0.243
11	0.737	32	0.412	52	0.237
12	0.717	33	0.401	53	0.230
13	0.697	34	0.390	54	0.224
14	0.678	35	0.379	55	0.218
15	0.660	36	0.369	56	0.212
16	0.642	37	0.359	57	0.206
17	0.624	38	0.349	58	0.200
18	0.607	39	0.339	59	0.195
19	0.591	40	0.330	60	0.189
20	0.574				

*Calibration time

To correct for physical decay of Rb 82, Table 6 shows the fraction of Rb 82 remaining in all 15 second intervals up to 300 seconds after time of calibration.

Table 6. Physical Decay Chart: Rb 82 half-life 75 seconds

Seconds	Fraction Remaining	Seconds	Fraction Remaining
0*	1	165	0.218
15	0.871	180	0.19
30	0.758	195	0.165
45	0.66	210	0.144
60	0.574	225	0.125
75	0.5	240	0.109
90	0.435	255	0.095
105	0.379	270	0.083
120	0.33	285	0.072
135	0.287	300	0.063
150	0.25		

*Elution time

12 CLINICAL PHARMACOLOGY

12.1 Mechanism of Action

Rb 82 is analogous to potassium ion (K^+) in its biochemical behavior and is extracted by the myocardium in proportion to the blood flow. Rb^+ participates in the sodium-potassium (Na^+/K^+) ion exchange pumps that are present in cell membranes. The intracellular uptake of Rb 82 requires maintenance of ionic gradient across cell membranes. Rb 82 radioactivity is increased in viable myocardium reflecting intracellular retention, while the tracer is cleared from necrotic or infarcted tissue.

12.2 Pharmacodynamics

In human studies, myocardial activity was noted within the first minute after peripheral intravenous injection of rubidium chloride Rb 82. When areas of infarction or ischemia are present in the myocardium, they are visualized within 2 minutes to 7 minutes after injection as photon-deficient, or “cold”, areas on the myocardial scan. In patients with reduced cardiac function, transit of the injected dose from the peripheral infusion site to the myocardium may be delayed [see *Dosage and Administration (2.3)*].

Blood flow brings Rb 82 to all areas of the body during the first pass of circulation. Accordingly, visible uptake is also observed in other highly vascularized organs, such as the kidneys, liver, spleen, and lungs.

12.3 Pharmacokinetics

With a physical half-life of 75 seconds, Rb 82 is converted by radioactive decay into stable Kr 82 gas, which is passively expired by the lungs.

13 NONCLINICAL TOXICOLOGY

13.1 Carcinogenesis, Mutagenesis, Impairment of Fertility

No long-term studies have been performed to evaluate carcinogenic potential, mutagenicity potential, or to determine whether rubidium chloride Rb 82 may affect fertility in males or females.

14 CLINICAL STUDIES

In a descriptive, prospective, blinded image interpretation study of adult patients with known or suspected coronary artery disease, myocardial perfusion deficits in stress and rest PET images obtained with ammonia N 13 (n = 111) or rubidium chloride Rb 82 (n = 82) were compared to changes in stenosis flow reserve (SFR) as determined by coronary angiography. PET perfusion defects at rest and stress for seven cardiac regions (anterior, apical, anteroseptal, posterolateral, anterolateral, posterolateral, and inferior walls) were graded on a scale of 0 (normal) to 5 (severe). Values for stenosis flow reserve, defined as flow at maximum coronary vasodilatation relative to rest flow, ranged from 0 (total occlusion) to 5 (normal). With increasing impairment of flow reserve, the subjective PET defect severity increased. A PET defect score of 2 or higher was positively correlated with flow reserve impairment (SFR<3).

A systematic review of published literature was conducted using pre-defined inclusion/exclusion criteria which resulted in identification of 10 studies evaluating the use of Rb 82 PET myocardial perfusion imaging (MPI) for the identification of coronary artery disease as defined by catheter-based angiography. In these studies, the patient was the unit of analysis and 50% stenosis was the threshold for clinically significant coronary artery disease (CAD). Of these 10 studies, 9 studies were included in a meta-analysis for sensitivity (excluding one study with 100% sensitivity) and 7 studies were included in a meta-analysis of specificity (excluding 3 studies with 100% specificity). A random effects model yielded overall estimates of sensitivity and specificity of 92% (95% CI: 89% to 95%) and 81% (95% CI: 76% to 86%), respectively. The use of meta-analysis in establishing performance characteristics is limited, particularly by the possibility of publication bias (positive results being more likely to be published than negative results) which is difficult to detect especially when based on a limited number of small studies.

16 HOW SUPPLIED/STORAGE AND HANDLING

How Supplied

CardioGen-82 (rubidium Rb 82 generator), used to produce Rubidium Chloride Rb 82 Injection, is supplied in a lead shield surrounded by a labeled plastic container with an activity at calibration time of 3,330 MBq to 5,550 MBq (90 mCi to 150 mCi) of Sr 82 adsorbed on a hydrous stannic oxide column (NDC 0270-0091-01).

The container label provides complete assay data for each generator.

Storage and Handling

Store the generator at 20°C to 25°C (68°F to 77°F) [See USP Controlled Room Temperature].

Receipt, transfer, possession, storage, disposal or use of this product is subject to the radioactive material regulations and licensing requirements of the U.S. Nuclear Regulatory Commission (NRC), Agreement States or Licensing States as appropriate. Do not dispose of the generator in regular refuse systems.

For questions about the disposal of the CardioGen-82 generator, contact Bracco Diagnostics Inc. at 1-800-447-6883, option 3.

17 PATIENT COUNSELING INFORMATION

Post-Study Voiding

Instruct patients to void after completion of each image acquisition session and as often as possible for one hour after completion of the PET scan [see *Warnings and Precautions (5.4)*].

Pregnancy

Advise a pregnant woman of the potential risk to a fetus [see *Use in Specific Populations (8.1)*].

Lactation

Advise lactating women that exposure to Rb 82 through breast milk can be minimized if breastfeeding is discontinued when Rubidium Chloride Rb 82 Injection is administered. Advise lactating women not to resume breastfeeding for at least one hour after administration of Rubidium Chloride Rb 82 Injection [see *Use in Specific Populations (8.2)*].

Manufactured for
Bracco Diagnostics Inc.

Princeton, NJ 08540

Manufactured By
GE Healthcare, Medi-Physics, Inc.
Arlington Heights, IL 60004

CardioGen-82[®] Rubidium 82 Generators, Infusion Systems, and Accessories contain proprietary technology covered by one or more patents listed at www.braccoimaging.com/us-en/patents.

CARDIOGEN-82[®] (Rubidium Rb 82 Generator)

Lot No.:

Sr-82 Activity: mBq mCi

Total Activity (Sr-85 + Sr-82): mBq mCi

Sr-85 + Sr-82 Ratio:

As of: Noon EST

Exp. Date: (see below)

Expiration Date:

Stop use of the CardioGen-82 generator once any one of the following Expiration Limits is reached.

- A total elution volume of 17 L has passed through the generator column, or
- 60 days post calibration date, or
- An eluate Sr-82 level of 0.01 microCi/mCi Rb-82, or
- An eluate Sr-85 level of 0.1 microCi/mCi Rb-82.

Manufactured for: Bracco Diagnostics Inc., Princeton, NJ 08540
by GE Healthcare, Medi-Physics, Inc., Arlington Heights, IL 60004

46-8200X